

FF (Florfenicol) Lateral Flow Assay Kit

Catalog No: E-FS-C040

20T/40T/80T

Version Number:	V1.2
Replace version:	V1.1
Revision Date:	2026.06.12

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Toll-free: 1-888-852-8623 Tel: 1-832-243-6086 Fax: 1-832-243-6017

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Test principle

This kit uses the principle of Immunochromatography assay for the qualitative detection. It can detect Florfenicol (FF) in samples, such as tissue, aquatic products, etc. After adding the sample solution into the sample well of detection card, FF of the sample solution combine with the gold-labelled antibody, so as to prevent the combining of gold-labelled antibody with FF conjugate on the cellulose membrane. When the concentration of FF in the sample solution is more than the detection limit, the detect line do not show color (or shows lighter color than control line) and the result is positive. When the concentration of FF in the sample solution is less than the detection limit, the detect line show color (shows equal or darker color than control line) and the result is negative.

Technical indicator

Detection limit: Tissue, aquatic products ----0.1 ppb.

Kits components

Item	Specifications
Detection card	20/40/80 T/kit
Gold-labelled micro well	20/40/80 wells/kit
Reconstitution Buffer	2/4/8 vials
Reagent A	2/4/8 vials
Manual	1 copy

Other materials required but not supplied

Instruments: Homogenizer, Nitrogen Evaporators, Water bath, Oscillators, Centrifuge, Graduated pipette, Balance (sensitivity 0.01 g)

High-precision transferpettor: Single channel (20-200 μ L, 100-1000 μ L)

Notes

1. FOR RESEARCH USE ONLY. Do not use product out of date or in a broken aluminum foil.
2. The detection card should be adjusted to room temperature after removed from the refrigerator before opening. The opening detection card should be used as soon as possible so as not to be invalid because of moisture.
3. Avoid of contacting the white membrane at the middle of the sample well.
4. The pipette cannot be mixing to avoid the cross-contaminant.
5. The tested sample should be clear, no turbidity particle and no bacterial pollution, otherwise it is easy to result in abnormal phenomena such as obstruction, unobvious color, etc., which affect the judgment of the experiment result.
6. If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.
7. The kit is used for rapid screening of actual samples. If the test result is positive, the instrument method such as HPLC, LC/MS, etc. can be used for quantitative confirmation.
8. Each reagent is optimized for use in the E-FS-C040. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other E-FS-C040 with different lot numbers.

Storage and expiry date

Storage: Store at 2-30°C. With cool and dry environment.

Expiry date: expiration date is on the packing box.

Sample pretreatment

Restore all reagents and samples to room temperature before use.

1. Sample pretreatment Notice:

Experimental apparatus should be clean, and the pipette should be disposable to avoid the experiment result be interfered by the contamination.

2. Sample pretreatment procedure:

2.1 Pretreatment of tissue, aquatic products sample:

- (1) Remove the skin and fat of sample, homogenize with homogenizer.
- (2) Weigh 2 ± 0.05 g of homogenized sample without fat into 15 mL centrifuge tube.
- (3) Add 2 mL of deionized water and 4 mL of **Reagent A**, tightly cap the vial and shake vigorously for 2 minutes. Centrifuge at 4000 r/minutes for 5 minutes at room temperature.
- (4) Transfer 1 mL of the upper layer solution into a 5 mL centrifuge tube, and dry it under a stream of nitrogen (or air) at 60°C. (Note: A small amount of unevaporated yellow oily residue at the bottom of the centrifuge tube is normal.)
- (5) Add 1 mL of **Reconstitution Buffer** to the dried centrifuge tube, mix thoroughly for testing (Testing should be carried out within 30 minutes).

Note: Detection limit: 0.1 ppb

Experiment procedure

1. Tear the aluminum foil bag of the detection card and take out the detection card, and put it on a smooth, clean table.
2. Take 100 μL (approx. 3 drops) of the test solution to the **Gold-labelled micro well**, gently pipette up and down for 30 seconds, whip the purple residual until it is completely dissolved (Avoid foaming), incubate at room temperature for 3 minutes.
3. Then gently pipette up and down for 10 seconds, remove all the liquid of the gold-labelled micro well into the sample well (S), count down at the same time.
4. Incubate for 5 to 8 minutes and then judge the results immediately.

Judgment of result

1. **Negative:** The control line region (C) show color, the test line region (T) shows equal or darker than line C. It indicates the content of FF in the sample is lower than detection limit or the sample doesn't contain FF.
2. **Positive:** The control line region (C) show color, the test line region (T) shows no color or lighter color than line C. It indicates the content of FF in the sample is higher than detection limit.
3. **Invalid:** The control line region (C) shows no color. It indicates operation process is wrong or the test card is invalid.

