

(FOR RESEARCH USE ONLY. DO NOT USE IT IN CLINICAL DIAGNOSIS !)

**Catalog No: E-BC-K892-M**

**Specification: 96T(40 samples)**

**Measuring instrument: Microplate reader (545-555 nm)**

**Detection range: 0.02-1 mmol/L**

## **Elabscience® Oxalate (Oxalic Acid) Colorimetric Assay Kit**

This manual must be read attentively and completely before using this product.

If you have any problem, please contact our Technical Service Center for help:

Toll-free: 1-888-852-8623

Tell: 1-832-243-6086

Fax: 1-832-243-6017

Email: [techsupport@elabscience.com](mailto:techsupport@elabscience.com)

Website: [www.elabscience.com](http://www.elabscience.com)

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

## Table of contents

<b>Assay summary .....</b>	<b>3</b>
<b>Intended use .....</b>	<b>4</b>
<b>Detection principle .....</b>	<b>4</b>
<b>Kit components &amp; storage .....</b>	<b>4</b>
<b>Materials prepared by users .....</b>	<b>5</b>
<b>Reagent preparation .....</b>	<b>5</b>
<b>Sample preparation .....</b>	<b>6</b>
<b>The key points of the assay .....</b>	<b>7</b>
<b>Operating steps .....</b>	<b>7</b>
<b>Calculation .....</b>	<b>8</b>
<b>Appendix I Performance Characteristics .....</b>	<b>9</b>
<b>Appendix II Example Analysis .....</b>	<b>11</b>
<b>Statement .....</b>	<b>12</b>

**Assay summary**



## Intended use

This kit can be used to measure oxalate content in serum, plasma and plant tissue samples.

## Detection principle

Oxalate oxidase catalyzes the oxidation of oxalate to produce hydrogen peroxide and carbon dioxide. Under the action of POD, hydrogen peroxide reacts with chromogenic substances to produce colored products. There is a specific absorption peak at 550 nm, and the color depth is proportional to the content of oxalate.

## Kit components & storage

Item	Component	Size (96 T)	Storage
Reagent 1	Working Solution	12 mL × 1 vial	-20℃, 12 months
Reagent 2	Chromogenic Agent A	12 mL × 1 vial	-20℃, 12 months shading light
Reagent 3	Chromogenic Agent B	12 mL × 1 vial	-20℃, 12 months shading light
Reagent 4	Enzyme Reagent A	Powder × 2 vials	-20℃, 12 months
Reagent 5	Enzyme Reagent B	Powder × 2 vials	-20℃, 12 months
Reagent 6	Regulator	12 mL × 1 vial	-20℃, 12 months
Reagent 7	Clarificant	2 mL × 1 vial	-20℃, 12 months
Reagent 8	1 mmol/L Standard	2 mL × 1 vial	-20℃, 12 months
	Microplate	96 wells	No requirement
	Plate Sealer	2 pieces	
	Sample Layout Sheet	1 piece	

Note: The reagents must be stored strictly according to the preservation conditions in the above table. The reagents in different kits cannot be mixed with each other. For a small volume of reagents, please centrifuge before use, so as not to obtain sufficient amount of reagents.

## Materials prepared by users

### Instruments:

Micropipettor, Microplate reader (545-555 nm, optimum wavelength: 550 nm), Centrifuge, 37°C incubator

### Reagents:

Double distilled water

## Reagent preparation

- ① Equilibrate all reagents to room temperature before use.
- ② Preparation of enzyme reagent A working solution:  
Dissolve one vial of enzyme reagent A with 5 mL of working solution, mix well to dissolve. Keep the prepared solution on ice during use.  
Store at 2-8°C for 2 days.
- ③ Preparation of enzyme reagent B working solution:  
Dissolve one vial of enzyme reagent B with 250 µL of double distilled water, mix well to dissolve. Store at 2-8°C for 2 days.
- ④ Preparation of reaction working solution:  
Before testing, please prepare sufficient reaction working solution according to the test wells. For example, prepare 201 µL of reaction working solution (mix well 100 µL of chromogenic agent A, 100 µL of chromogenic agent B and 1 µL of enzyme reagent B working solution). The prepared solution should be prepared on spot and used up within 1 h.

⑤ The preparation of standard curve:

Always prepare a fresh set of standards. Discard working standard dilutions after use.

Dilute 1 mmol/L standard with double distilled water to a serial concentration. The recommended dilution gradient is as follows: 0, 0.2, 0.4, 0.5, 0.6, 0.7, 0.8, 1.0 mmol/L. Reference is as follows:

Item	①	②	③	④	⑤	⑥	⑦	⑧
<b>Concentration (mmol/L)</b>	<b>0</b>	<b>0.2</b>	<b>0.4</b>	<b>0.5</b>	<b>0.6</b>	<b>0.7</b>	<b>0.8</b>	<b>1.0</b>
<b>1 mmol/L standard (μL)</b>	0	40	80	100	120	140	160	200
<b>Double distilled water (μL)</b>	200	160	120	100	80	60	40	0

## Sample preparation

### ① Sample preparation

**Serum and plasma:** detect directly. If not detected on the same day, the serum or plasma can be stored at -80℃ for a month.

### **Plant tissue sample:**

- ① Harvest the amount of plant tissue needed for each assay (initial recommendation 20 mg).
- ② Homogenize 20 mg tissue in 180 μL double distilled water with a dounce homogenizer at 4℃.
- ③ Centrifuge at 10000×g for 10 min at 4℃ to remove insoluble material. Collect supernatant.
- ④ For each well, add 10 μL of supernatant and 10 μL of regulator, mix well. Stand at room temperature for 10 min for detection.

## ② Dilution of sample

The recommended dilution factor for different samples is as follows (for reference only):

Sample type	Dilution factor
Human plasma	1
10% <i>Epipremnum aureum</i> tissue homogenate	2-3
Rats plasma	1

Note: The diluent is double distilled water. For the dilution of other sample types, please do pretest to confirm the dilution factor.

## The key points of the assay

- ① Adjust the dilution ratio of samples according to the pre-experiment results.
- ② Pay attention to the storage conditions when using enzyme reagents..

## Operating steps

- ① Sample well: Add 10  $\mu\text{L}$  of sample to corresponding well.  
Control well: Add 10  $\mu\text{L}$  of sample to corresponding well.  
Standard well: Add 10  $\mu\text{L}$  of standard with different concentrations to corresponding well.
- ② Add 80  $\mu\text{L}$  of enzyme reagent A working solution into sample well and standard well. Add 80  $\mu\text{L}$  of double distilled water into control well.
- ③ Incubate at 37°C for 10 min.
- ④ Add 120  $\mu\text{L}$  of reaction working solution into each well.
- ⑤ Stand at room temperature for 2 min.
- ⑥ Add 20  $\mu\text{L}$  of clarificant into each well.
- ⑦ Mix fully with microplate reader 5 s and measure the OD value of each well at 550 nm with microplate reader.

## Calculation

### The standard curve:

1. Average the duplicate reading for each standard.
2. Subtract the mean OD value of the blank (Standard # ① ) from all standard readings. This is the absolved OD value.
3. Plot the standard curve by using absolved OD value of standard and correspondent concentration as y-axis and x-axis respectively. Create the standard curve ( $y = ax + b$ ) with graph software (or EXCEL).

### The sample:

#### 1. Serum/plasma sample:

$$\text{Oxalate content (mmol/L)} = (\Delta A - b) \div a \times f$$

#### 2. Plant tissue sample:

$$\text{Oxalate content (mmol/ kg wet weight)} = (\Delta A - b) \div a \div (m \div V) \times 2 \times f$$

### [Note]

$$\Delta A: OD_{\text{Sample}} - OD_{\text{Control}}$$

2\*: Dilution factor of sample in sample pretreatment step.

m: the weight of sample, 0.1 g.

V: the volume of sample homogenate, 0.9 mL.

f: Dilution factor of sample before tested



## Appendix I Performance Characteristics

### 1. Parameter:

#### Intra-assay Precision

Three human serum samples were assayed in replicates of 20 to determine precision within an assay (CV = Coefficient of Variation).

Parameters	Sample 1	Sample 2	Sample 3
Mean (mmol/L)	0.08	0.54	0.88
%CV	3.5	3.1	2.4

#### Inter-assay Precision

Three human serum samples were assayed 20 times in duplicate by three operators to determine precision between assays.

Parameters	Sample 1	Sample 2	Sample 3
Mean (mmol/L)	0.08	0.54	0.88
%CV	4.8	5.2	4.1

#### Recovery

Take three samples of high concentration, middle concentration and low concentration to test the samples of each concentration for 6 times parallelly to get the average recovery rate of 95%.

	Standard 1	Standard 2	Standard 3
Expected Conc. (mmol/L)	0.35	0.55	0.75
Observed Conc. (mmol/L)	0.3	0.5	0.7
Recovery rate (%)	96	94	95

#### Sensitivity


The analytical sensitivity of the assay is 0.02 mmol/L. This was determined by adding two standard deviations to the mean O.D. obtained when the zero standard was assayed 20 times, and calculating the

corresponding concentration.

## 2. Standard curve:

As the OD value of the standard curve may vary according to the conditions of the actual assay performance (e.g. operator, pipetting technique or temperature effects), so the standard curve and data are provided as below for reference only:

Concentration (mmol/L)	0	0.2	0.4	0.5	0.6	0.7	0.8	1.0
Average OD	0.046	0.137	0.228	0.268	0.316	0.353	0.390	0.468
Absoluted OD	0.000	0.090	0.182	0.222	0.270	0.307	0.344	0.422



## Appendix Π Example Analysis


### Example analysis :

For epipremnum aureum, take 10  $\mu\text{L}$  of 10% epipremnum aureum tissue homogenate and carry the assay according to the operation table. The results are as follows:

Standard curve:  $y = 0.4236x + 0.0072$ , the average OD value of the sample is 0.094, the average OD value of the control is 0.053, and the calculation result is:

$$\begin{aligned}\text{Oxalate content (mmol/kg wet weight)} &= (0.094 - 0.053 - 0.0072) \div 0.4236 \times 2 \times 9 \\ &= 1.43 \text{ mmol/kg wet weight}\end{aligned}$$

Detect rats plasma, human plasma, 10% epipremnum aureum tissue homogenate according to the protocol, the result is as follows:



## Statement

1. This assay kit is for Research Use Only. We will not response for any arising problems or legal responsibilities causing by using the kit for clinical diagnosis or other purpose.
2. Please read the instructions carefully and adjust the instruments before the experiments. Please follow the instructions strictly during the experiments.
3. Protection methods must be taken by wearing lab coat and latex gloves.
4. If the concentration of substance is not within the detection range exactly, an extra dilution or concentration should be taken for the sample.
5. It is recommended to take a pre-test if your sample is not listed in the instruction book.
6. The experimental results are closely related to the situation of reagents, operations, environment and so on. Elabscience will guarantee the quality of the kits only, and NOT be responsible for the sample consumption caused by using the assay kits. It is better to calculate the possible usage of sample and reserve sufficient samples before use.