(FOR RESEARCH USE ONLY, DO NOT USE IT IN CLINICAL DIAGNOSIS!)

Catalog No: E-BC-K235-S

Specification: 50 Assays (19 samples)/100 Assays(44 samples)

Measuring instrument: Spectrophotometer (505 nm)

Detection range: 1.26-72.3 IU/L

Elabscience® Alanine Aminotransferase (ALT/GPT) Activity Assay Kit

This manual must be read attentively and completely before using this product.

If you have any problem, please contact our Technical Service Center for help:

Toll-free: 1-888-852-8623

Tell: 1-832-243-6086

Fax: 1-832-243-6017

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Table of contents

| Assay summary | 3 |
|--|----|
| Intended use | 4 |
| Detection principle | 4 |
| Kit components & storage | 5 |
| Materials prepared by users | 5 |
| Reagent preparation | 6 |
| Sample preparation | 6 |
| The key points of the assay | 7 |
| Operating steps | 8 |
| Calculation | 10 |
| Appendix I Performance Characteristics | 11 |
| Appendix П Example Analysis | 13 |
| Appendix III Publications | 14 |
| Statement | 15 |

Assay summary



Intended use

This kit can be used to measure ALT/GPT activity in serum, plasma, tissue, culture supernatant and other samples.

Detection principle

Alanine aminotransferase (ALT) catalyze the amino conversion reaction between alanine and α -ketoglutaric acid to produce pyruvic acid and glutamic acid at pH 7.4 and 37°C. Then phenylhydrazine was added to form phenylhydrazone with pyruvic acid. Phenylhydrazone is reddish brown under alkaline conditions. ALT activity can be calculated by measuring the OD values at 510 nm.

Kit components & storage

| Item | Component | Size 1 (50 assays) | Size 2 (100 assays) | Storage | |
|-----------|-----------------------------|-----------------------|------------------------|--------------------------------------|--|
| Reagent 1 | Buffer Solution | 1.8 mL × 1 vial | 1.8 mL × 1 vial | 2-8°C, 12 months | |
| Reagent 2 | 2 mmol/L Sodium Pyruvate | 1.8 mL × 1 vial | 1.8 mL × 1 vial | 2-8°C, 12 months | |
| Reagent 3 | Substrate Solution | 30 mL × 1 vial | 30 mL × 2 vials | 2-8°C, 12 months | |
| Reagent 4 | Chromogenic Agent | 30 mL × 1 vial | 30 mL × 2 vials | 2-8°C, 12 months shading light | |
| Reagent 5 | Alkali Reagent | 30 mL × 1 vial | 30 mL × 2 vials | 2-8°C, 12 months | |

Note: The reagents must be stored strictly according to the preservation conditions in the above table. The reagents in different kits cannot be mixed with each other. For a small volume of reagents, please centrifuge before use, so as not to obtain sufficient amount of reagents.

Materials prepared by users

Instruments:

Spectrophotometer (505 nm), Micropipettor, Vortex mixer, Incubator

Reagents:

Double distilled water, Normal saline (0.9% NaCl), PBS (0.01 M, pH 7.4)

Reagent preparation

- ① The preparation of alkali working solution:

 For each tube, prepare 5 mL of alkali working solution (mix well 0.5 mL of alkali reagent and 4.5 mL of double distilled water). The alkali working solution should be prepared on spot.
- 2 Incubate substrate solution at 37 °C for 10 min.

Sample preparation

① Sample preparation

Serum and plasma: detect directly. If not detected on the same day, the serum or plasma can be stored at -80°C for a month.

Urine: collect fresh urine and centrifuge at 10000×g for 15 min at 4°C. Take the supernatant and preserve it on ice for detection.

Tissue sample:

- ① Harvest the amount of tissue needed for each assay (initial recommendation 20 mg).
- 2 Wash tissue in cold PBS (0.01 M, pH 7.4).
- \odot Homogenize 20 mg tissue in 180 μ L PBS (0.01 M, pH 7.4) with a dounce homogenizer at 4°C.
- ④ Centrifuge at 10000×g for 10 minutes to remove insoluble material.
 Collect supernatant and keep it on ice for detection.
- (E-BC-K318-M).

Cells:

- ① Harvest the number of cells needed for each assay (initial recommendation 1×10⁶ cells).
- ② Wash cells with PBS (0.01 M, pH 7.4).
- \odot Homogenize 1×10⁶ cells in 300 µL PBS (0.01M, pH 7.4) with a ultrasonic cell disruptor at 4°C.
- ④ Centrifuge at 10000×g for 10 minutes to remove insoluble material.
 Collect supernatant and keep it on ice for detection.
- (E-BC-K318-M).

2 Dilution of sample

The recommended dilution factor for different samples is as follows (for reference only):

| Sample type | Dilution factor |
|--------------------------------------|-----------------|
| 10% Rat liver tissue homogenization | 30-60 |
| 10% Rat kidney tissue homogenization | 1 |
| Human serum | 1 |
| Human plasma | 1 |
| HepG2 cells homogenization | 1 |

Note: The diluent is saline (0.9% NaCl) or PBS (0.01 M, pH 7.4). For the dilution of other sample types, please do pretest to confirm the dilution factor

The key points of the assay

Detect the sample as soon as possible after collection. The serum sample can be store at 2-8°C for 7 days and -20°C for 20 days.

Operating steps

The measurement of standard curve

- ① Standard tubes: Record the test tube with A, B, C, D, E, F in duplication, add 0.1 mL of buffer solution to the standard tubes respectively. Add 0, 0.05, 0.10, 0.15, 0.20, 0.25 mL of 2 mmol/L sodium pyruvate to the standard tubes from A to F, respectively. Add 0.50, 0.45, 0.40, 0.35, 0.30, 0.25 mL of substrate solution to the standard tubes from A to F, respectively.
- ② Add 0.50 mL of chromogenic agent to each tube.
- ③ Mix fully and incubate at 37°C for 20 min.
- ④ Add 5 mL of alkali working solution to each tube.
- ⑤ Stand for 10 min at room temperature and set to zero with double distilled water and measure the OD value of each tube with 1 cm optical path cuvette at 505 nm.

Note: Plot the standard curve by using absoluted OD value(OD value of each tube - OD value of A tube) of standard and correspondent concentration (0, 28, 57,97, 150, 200) as x-axis and y-axis respectively. Create the standard curve (y =ax²+bx + c) with graph software (or EXCEL).

The measurement of samples

- Sample tubes: add 0.5 mL of substrate solution (pre-heated at 37°C for 10 min) and 0.1 mL of sample.
 Control tubes: add 0.5mL of substrate solution (pre-heated at 37°C for 10 min).
- ② Mix well and incubate at 37°C for 30 min.
- ③ Add 0.50 mL of chromogenic agent to each tube.
- ④ Control tubes: Add 0.1 mL of sample to control tubes.
- ⑤ Mix well and incubate at 37°C for 20 min.
- 6 Add 5 mL of alkali working solution to each tube.
- Stand for 10 min at room temperature and set to zero with double distilled water and measure the OD value of each tube with 1 cm optical path cuvette at 505 nm.

Calculation

The standard curve:

- 1. Average the duplicate reading for each standard.
- 2. Subtract the mean OD value of the blank (Standard #A) from all standard readings. This is the absoluted OD value.
- 3. Plot the standard curve by using absoluted OD value of standard and correspondent concentration (0, 28, 57, 97, 150, 200) as x-axis and y-axis respectively. Create the standard curve ($\mathbf{y} = \mathbf{a}\mathbf{x}^2 + \mathbf{b}\mathbf{x} + \mathbf{c}$) with graph software (or EXCEL).

Definition of international unit: The enzyme amount of 1 μ mol of NADH consumed in reaction system (1 mL sample/1 g tissue protein, 25°C) per minute is defined as 1 unit (wavelength is 340 nm, optical path is 1 cm).

Definition of carmen unit: 1 mL of sample, the total volume of reaction is 3 mL, wavelength is 340 nm, optical path is 1 cm, react at 25° C for 1 min, the amount of generated pyruvic acid which oxidize NADH to NAD⁺ and cause absorbance decreasing 0.001 is as 1 unit. (1 carmen unit = 0.482 IU/L, 25° C).

The sample:

1. Serum (plasma) sample:

ALT/GPT activity (IU/L) = $[a \times (\Delta A_{505})^2 + b \times \Delta A_{510} + c] \times 0.482 \times f$

2. Tissue and cells sample:

ALT/GPT activity (IU/gprot) = $[a \times (\Delta A_{505})^2 + b \times \Delta A_{510} + c] \times 0.482 \times f \div C_{pr}$

[Note]

ΔA₅₀₅: OD_{sample} - OD_{control}

f: Dilution factor of sample before tested.

C_{pr}: Concentration of protein in sample (gprot/L).

Appendix I Performance Characteristics

1. Parameter:

Intra-assay Precision

Three human serum samples were assayed in replicates of 20 to determine precision within an assay. (CV = Coefficient of Variation)

| Parameters Sample 1 | | Sample 2 | Sample 3 | |
|---------------------|-------------|----------|----------|--|
| Mean (IU/L) 5.50 | | 35.80 | 62.50 | |
| %CV | 4.6 4.2 4.1 | | 4.1 | |

Inter-assay Precision

Three human serum samples were assayed 17 times in duplicate by three operators to determine precision between assays.

| Parameters | Sample 1 | Sample 2 | Sample 3 |
|-------------|----------|----------|----------|
| Mean (IU/L) | 5.50 | 35.80 | 62.50 |
| %CV | 8.0 | 8.4 | 8.2 |


Sensitivity

The analytical sensitivity of the assay is 1.26 IU/L. This was determined by adding two standard deviations to the mean O.D. obtained when the zero standard was assayed 20 times, and calculating the corresponding concentration.

2. Standard curve:

As the OD value of the standard curve may vary according to the conditions of the actual assay performance (e.g. operator, pipetting technique or temperature effects), so the standard curve and data are provided as below for reference only:

| Carmen unit | 0 | 28 | 57 | 97 | 150 | 200 |
|--------------|-------|-------|-------|-------|-------|-------|
| Average OD | 0.274 | 0.387 | 0.497 | 0.585 | 0.674 | 0.750 |
| Absoluted OD | 0 | 0.113 | 0.224 | 0.312 | 0.401 | 0.476 |



Appendix Π Example Analysis


Example analysis:

Take 0.1 mL of human serum, carry the assay according to the operation steps. The results are as follows:

standard curve: $y = 629.84 x^2 + 115.13 x + 2.0232$, the average OD value of the sample is 0.288, the average OD value of the control is 0.240, and the calculation result is:

ALT activity
$$(IU/L)$$
 = [629.84× (0.288 - 0.240)²+ 115.13 × (0.288 - 0.240) +2.0232] × 0.482
= 4.34 IU/L

Detect human serum (V=0.1 mL), 10% rat liver tissue homogenate (the concentration of protein is 13.11 gprot/L, dilute for 50 times, V= 0.1 mL), 10% rat kidney tissue homogenate (the concentration of protein is 9.26 gprot/L, V= 0.1 mL), HepG2 cells (the concentration of protein is 1.83 gprot/L, V= 0.1 mL) according to the protocol, the result is as follows:



Appendix III Publications

- 1. Fang X, Ding H, Chen Y, et al. Wireless Optogenetic Targeting Nociceptors Helps Host Cells Win the Competitive Colonization in Implant Associated Infections[J]. Small Methods, 2024, 8(12): 2400216.
- 2. Lyu Y, Yang X, Yang L, et al. Lipid nanoparticle-mediated hepatocyte delivery of siRNA and silibinin in metabolic dysfunction-associated steatotic liver disease[J]. Journal of Controlled Release, 2024, 373: 385-398.
- 3. Li C, Wu Y, Chen K, et al. Gp78 deficiency in hepatocytes alleviates hepatic ischemia-reperfusion injury via suppressing ACSL4-mediated ferroptosis[J]. Cell Death & Disease, 2023, 14(12): 810.
- 4. Liu D, Zhan J, Wang S, et al. Chrysanthemum morifolium attenuates metabolic and alcohol-associated liver disease via gut microbiota and PPARa/y activation[J]. Phytomedicine, 2024, 130: 155774.
- 5. Ma Y, Zhou W, Wang H, et al. The double-layer emulsions loaded with bitter melon (Momordica charantia L.) seed oil protect against dextran sulfate sodium-induced ulcerative colitis in mice[J]. International Journal of Biological Macromolecules, 2024, 278: 134279.
- 6. Wang Y, Wang J, Zhou T, et al. Investigating the potential mechanism and therapeutic effects of SLXG for cholesterol gallstone treatment[J]. Phytomedicine, 2024, 132: 155886.

Statement

- 1. This assay kit is for Research Use Only. We will not response for any arising problems or legal responsibilities causing by using the kit for clinical diagnosis or other purpose.
- Please read the instructions carefully and adjust the instruments before the experiments. Please follow the instructions strictly during the experiments.
- 3. Protection methods must be taken by wearing lab coat and latex gloves.
- 4. If the concentration of substance is not within the detection range exactly, an extra dilution or concentration should be taken for the sample.
- 5. It is recommended to take a pre-test if your sample is not listed in the instruction book.
- 6. The experimental results are closely related to the situation of reagents, operations, environment and so on. Elabscience will guarantee the quality of the kits only, and NOT be responsible for the sample consumption caused by using the assay kits. It is better to calculate the possible usage of sample and reserve sufficient samples before use.