

Urease Test Kit

Catalog No: E-IA-C078

100T

Version Number: V1.0
Replace version: /
Revision Date: 2026.05.20

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Toll-free: 1-888-852-8623 Tel: 1-832-243-6086 Fax: 1-832-243-6017

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Test principle

This product is used for the rapid detection of Urease in Soymilk sample, featuring simple operation and short detection time.

Technical indicator

Detection limit: 0.2U.

Kits components

Item	Specifications
Detection Solution A	1 vial
Detection Solution B	1 vial
1.5 mL Test Tube	1 pack
Disposable Dropper	1 pack
Color Comparison Card	1 piece
Manual	1 copy

Note: All reagent bottle caps must be tightened to prevent evaporation and microbial pollution.

Other materials required but not supplied

Instruments: Homogenizer, Nitrogen Evaporators, Water bath, Centrifuge, Graduated pipette, Balance (sensitivity 0.01g), Oscillators, EP tubes.

Micropipette: Single channel (20-200 µL, 100-1000 µL)

Notes

- FOR RESEARCH USE ONLY. Do not use product out of date.
- The test tubes can be reused after rinsing and air-drying.
- Avoid contact of **Detection Solution B** with skin and eyes during detection to prevent corrosion.
- When it is difficult to judge the sample, use boiled soymilk for control test.
- Shake the detection solution well before use and add dropwise, store away from light at all times.
- Each reagent is optimized for use in the E-IA-C078. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other E-IA-C078 with different lot numbers.**
- Dispose of solid and liquid waste generated after using this product in accordance with local regulatory requirements to prevent contamination.
- This method is applicable to on-site rapid testing. Samples whose test results do not comply with standard limits or labeled values shall be measured in triplicate.
- The kit is used for rapid screening of actual samples. If the test result is positive, the instrument method such as HPLC, LC/MS, etc. can be used for quantitative confirmation.

Storage and expiry date

Storage: Store at 2-30°C. With cool and dry environment.

Expiry date: expiration date is on the packing box.

Operating Procedures

1. Take 1 mL of soymilk into a 1.5 mL test tube, add 3 drops of **Detection Solution A**, cap and shake well.
2. After 1 to 2 minutes, add 1 drop of **Detection Solution B**, cap the tube and shake well. Observe the color of the solution after 10 minutes.

Judgment of result

1. **Negative:** Light yellow (original color), consistent with the sample color before detection.
Positive: If the color differs from the original sample color, compare it with the **Color Comparison Card**. The closest color indicates the urease content.