

Sorbic Acid Test Kit

Catalog No: E-IA-C052

100T

Version Number:	V1.0
Replace version:	/
Revision Date:	2026.09.22

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Toll-free: 1-888-852-8623 Tel: 1-832-243-6086 Fax: 1-832-243-6017

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Test principle

This product is used for the rapid detection of Sorbic Acid in food, beverages, etc, featuring simple operation and short detection time.

Kits components

Item	Specifications
Detection Solution A	2 vials
Detection Solution B	2 vials
10 mL Test Tube	1 pack
2 mL Test Tube	1 pack
Dropper	1 pack
Filter paper	1 pack
Color Comparison Card	1 piece
Manual	1 copy

Note: All reagent bottle caps must be tightened to prevent evaporation and microbial pollution.

Other materials required but not supplied

Instruments: Homogenizer, Nitrogen Evaporators, Water bath, Centrifuge, Graduated pipette, Balance (sensitivity 0.01g), Oscillators, EP tubes.

Micropipette: Single channel (20-200 µL, 100-1000 µL)

Notes

1. FOR RESEARCH USE ONLY. Do not use product out of date.
2. The test tubes can be reused after rinsing and air-drying.
3. Shake the detection solution well before use and add dropwise, store away from light at all times.
4. **Each reagent is optimized for use in the E-IA-C052. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other E-IA-C052 with different lot numbers.**
5. Dispose of solid and liquid waste generated after using this product in accordance with local regulatory requirements to prevent contamination.
6. This method is applicable to on-site rapid testing. Samples whose test results do not comply with standard limits or labeled values shall be measured in triplicate.
7. The kit is used for rapid screening of actual samples. If the test result is positive, the instrument method such as HPLC, LC/MS, etc. can be used for quantitative confirmation.

Storage and expiry date

Storage: Store at 2-30°C. With cool and dry environment.

Expiry date: expiration date is on the packing box.

Operating Procedures

1. Weigh $1.0 \text{ g} \pm 0.05 \text{ g}$ (mL) of sample into a **10 mL Test Tube**, add distilled water up to the 5 mL mark, sonicate for 5 min, filter, and set aside for later use.
2. Transfer 1 mL of sample solution into a **2 mL Test Tube**, add 3 drops of **Detection Solution A**, heat in a water bath or metal bath at 90°C for 10 min.
3. Add 3 drops of **Detection Solution B**, heat in a water bath or metal bath at 90°C for 10 min.
4. Observe the color and compare it with the **Color Comparison Card**.

Judgment of result

Compare the results with the **Color Comparison Card**.