

ERY (Erythromycin) ELISA Kit

Catalog No: E-FS-E083

96T/96T*3

Version Number:	V1.2
Replace version:	V1.1
Revision Date:	2025.08.26

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Toll-free: 1-888-852-8623 Tel: 1-832-243-6086 Fax: 1-832-243-6017

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Test principle

This kit uses Competitive-ELISA as the method for the quantitative detection. It can detect Erythromycin (ERY) in samples, such as muscle, milk, liver etc. This kit is composed of ELISA Microtiter plate, HRP conjugate, antibody working solution, standard and other supplementary reagents. The microtiter plate in this kit has been pre-coated with coupled antigen. During the reaction, ERY in the samples or standard competes with coupled antigen on the solid phase supporter for sites of anti-ERY antibody. Then Horseradish Peroxidase (HRP) conjugate is added to each microtiter plate well, and substrate reagent is added for color development. There is a negative correlation between the OD value of samples and the concentration of ERY. The concentration of ERY in the samples can be calculated by comparing the OD of the samples to the standard curve.

Technical indicator

Reaction mode (Incubation time and temperature): 25°C, 30 min; 15 min.

Detection limit: Raw milk, Egg, Muscle, Liver, Fish, Shrimp ---5 ppb; Pig urine ---30 ppb.

Cross-reactivity: Erythromycin ----100%; Spiramycin, Valnemulin, Lincomycin, Kanamycin, Tilmicosin, Doramectin, Ivermectin, Eprinomectin, Moxidectin, Abamectin, Tylosin---<0.1%

Sample recovery rate: 90%±30%.

Kits components

Item	Specifications
ELISA Microtiter Plate	96 wells
Standard Liquid	1.5 mL each (ppb=ng/mL=ng/g) (0 ppb, 0.1 ppb, 0.3 ppb, 0.9 ppb, 2.7 ppb, 8.1 ppb)
HRP Conjugate	7 mL
Antibody Working Solution	7 mL
Substrate Reagent A	7 mL
Substrate Reagent B	7 mL
Stop Solution	7 mL
20×Concentrated Wash Buffer	25 mL
20×Concentrated Sample Diluent	25 mL
Milk Diluent	50 mL
Plate Sealer	3 pieces
Sealed Bag	1 piece
Manual	1 copy

Note: All reagent bottle caps must be tightened to prevent evaporation and microbial pollution.

Other materials required but not supplied

Instrument: Microplate reader, Vortex mixer, Centrifuge, Balance (sensitivity 0.01 g).

Micropipette: Single channel (20-200 µL, 100-1000 µL), Multichannel (30-300 µL).

Reagents: Methanol, HCl.

Notes

1. The overall OD value will be lower when reagents have not been brought to room temperature before use or room temperature is below 25°C.
2. If the wells turn dry during the washing procedure, it will lead to bad linear standard curve and poor repeatability. Operate the next step immediately after wash.
3. Mix thoroughly and wash the plate completely. The consistency of wash procedure can strongly affect the reproducibility of this ELISA kit.
4. FOR RESEARCH USE ONLY. ELISA Microtiter plate should be covered by plate sealer. Avoid the kit to strong light.
5. **Each reagent is optimized for use in the E-FS-E083. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other E-FS-E083 with different lot numbers.**
6. Substrate Reagent should be abandoned if it turns blue color. When OD value of standard (concentration: 0) < 0.8 unit (A450nm < 0.8), it indicates the reagent may be deteriorated.
7. Stop solution is caustic, avoid contact with skin and eyes.
8. As the OD values of the standard curve may vary according to the conditions of the actual assay performance (e.g. operator, pipetting technique, washing technique or temperature effects), the operator should establish a standard curve for each test.
9. Even the same operator might get different results in two separate experiments. In order to get reproducible results, the operation of every step in the assay should be controlled.
10. **For mentioned sample fast and efficient extraction methods are included in the kit description. Please consult technical support for the applicability if other sample need to be tested.**
11. The kit is used for rapid screening of actual samples. If the test result is positive, the instrument method such as HPLC, LC/MS, etc. can be used for quantitative confirmation.

Storage and expiry date

Store the kit at 2-8°C. Do not freeze any test kit components.

Return any unused microwells to their original foil bag and reseal them together with the desiccant provided and further store at 2-8°C.

Expiry date: expiration date is on the packing box.

Experimental preparation

Restore all reagents and samples to room temperature before use.

Open the microplate reader in advance, preheat the instrument, and set the testing parameters

1. Sample pretreatment Notice:

Experimental apparatus should be clean, and the pipette should be disposable to avoid cross-contamination during the experiment.

2. Solution preparation

Please prepare solution according to the number of samples. Don't use up all components in the kit at once!

Solution 1: 0.1 M HCl Solution (*for raw milk sample*)

Dilute 0.86 mL of **HCl** to 100 mL with deionized water, mix fully.

Solution 2: Wash Buffer

Dilute **20×Concentrated Wash Buffer** with deionized water. (**20×Concentrated Wash Buffer** (V): Deionized water (V) = 1:19).

Solution 3: Sample Diluent (*for livestock, egg, fish, shrimp sample*)

Dilute **20×Concentrated Sample Diluent** with deionized water. (**20×Concentrated Sample Diluent** (V): Deionized water (V) = 1:19).

3. Sample pretreatment procedure

3.1 Pretreatment of raw milk sample:

- (1) Take 1 mL of the fresh raw milk into a centrifuge tube, and add 100 μ L of **0.1 M HCl Solution** (Solution 1). Vortex for 1 min to mix thoroughly. Centrifuge for 10 min at 4000 g.
- (2) Remove the upper layer of fat, take 50 μ L of the solution in intermediate layer into another centrifuge tube, and add 450 μ L of **Milk Diluent**. Vortex for 10 s to mix thoroughly.
- (3) Take 50 μ L for detection.

Note: Sample dilution factor: 10, detection limit: 5 ppb

3.2 Pretreatment of muscle (livestock, fish, shrimp), egg, liver sample:

- (1) Remove fat from sample. Homogenize the representative sample with a homogenizer and mix fully.
- (2) Weigh 1 ± 0.05 g of homogenate sample into a 50 mL of centrifuge tube. Add 0.25 mL of **Methanol** and 4.75 mL of **Sample Diluent** (Solution 3), and vortex for 1 min.
- (3) Centrifuge for 10 min at 4000 g.
- (4) Take 50 μ L for analysis

Note: Samples that are beyond the detection range should be diluted again with 3% NaCl by a factor of 5 to 10.

Note: Sample dilution factor: 5, detection limit: 5 ppb

3.3 Pretreatment of pig urine sample:

- (1) Take 20 μ L of the sample into a centrifuge tube, and add 1980 μ L of **Sample Diluent** (Solution 3).

Vortex for 10s to mix thoroughly.

(2) Take 50 µL for detection.

Note: Sample dilution factor: 100, detection limit: 30 ppb

Assay procedure

Restore all reagents and samples to room temperature (25°C) before use. All the reagents should be mixed thoroughly by gently swirling before pipetting. Avoid foaming. The unused ELISA Microtiter plate should be sealed as soon as possible and stored at 2-8°C.

1. **Number:** number the sample and standard in order (multiple well), and keep a record of standard wells and sample wells. **Standard and Samples need test in duplicate.**
2. **Add Sample:** add 50 µL of **Standard or Sample** per well, then add 50 µL **HRP Conjugate** to each well. Add 50 µL **Antibody Working Solution**, cover the plate with plate sealer, and oscillate for 10 s gently to mix thoroughly. Incubation for 30 min at 25°C in shading light.
3. **Wash:** uncover the sealer carefully, remove the liquid in each well. Immediately add 260 µL of **Wash Buffer** (Solution 2) to each well and wash. Repeat wash procedure for 4 times, 30 s intervals/time. Invert the plate and pat it against thick clean absorbent paper (If bubbles exist in the wells, clean tips can be used to prick them).
4. **Color Development:** add 100 µL of mixture of **Substrate Reagent A** and **Substrate Reagent B**. Gently oscillate for 10s to mix thoroughly. Incubate at 25°C for 15-20 min in shading light (The reaction time can be extended according to the actual color change).
Note: Substrate Reagent A and Substrate Reagent B solution are mixed in a volume ratio of 1:1. It must be thoroughly mixed. The mixture should be used within 5 min.
5. **Stop reaction:** add 50 µL of **Stop Solution** to each well, gently oscillate for 10 s and mix fully to stop the reaction.
6. **OD Measurement:** determine the optical density (OD value) of each well at 450 nm (reference wavelength 630 nm) with a microplate reader. This step should be finished in 5 min after stop reaction.

Result analysis

1. Absorbance (%)= $A/A_0 \times 100\%$

A: Average absorbance of standard or sample

A_0 : Average absorbance of 0 ppb Standard

2. Drawing and calculation of standard curve

Create a standard curve by plotting the absorbance percentage of each standard on the y-axis against the log concentration on the x-axis to draw a semi-logarithmic plot. Add average absorbance value of sample to standard curve to get corresponding concentration. **If samples have been diluted, the concentration calculated from the standard curve must be multiplied by the dilution factor.**

For this kit, it is more convenient to use professional analysis form for accurate and fast analysis of batch samples.

Erythromycin (E-FS-E083) Standard Curve

