

**SDS-PAGE Gel kit**

Cat. No: E-IR-R305

Size: 10 Assays/ 25 Assays / 50 Assays

| Cat.          | Products           | 10 Assays | 25 Assays       | 50 Assays | Storage |
|---------------|--------------------|-----------|-----------------|-----------|---------|
| E-IR-R305A    | 30% Acr-Bis (29:1) | 45 mL     | 100 mL          | 100 mLx2  | 2~8°C   |
| E-IR-R305B    | Separating Gel Mix | 22 mL     | 55 mL           | 55 mL x2  | RT      |
| E-IR-R305C    | Stacking Gel Mix   | 6 mL      | 12 mL           | 22 mL     | RT      |
| E-IR-R305D    | APS                | 170 mg    | 170 mg x 2      | 700 mg    | RT      |
| E-IR-R305E    | TEMED              | 100 µL    | 250 µL          | 500 µL    | 2~8°C   |
| <b>Manual</b> |                    |           | <b>One Copy</b> |           |         |

**Introduction**

Elabscience® SDS-PAGE Gel Assay kit is the classic SDS-PAGE gel preparation kit which contains all kinds of reagents needed to prepare the SDS-PAGE gel. Users only need apparatus and ddH<sub>2</sub>O to complete the gel preparation.

The Gel Mix in this kit is the mixture of the gel buffer such as SDS and Tris-HCl and so on, which simplify the procedure of gel preparation. There are Tris-HCl pH 8.8 and SDS in the Separating Gel Mix, Tris-HCl pH 6.8 and SDS are in the Stacking Gel Mix.

**Instructions**

According to the different molecular weight of the target protein, different concentration of separating gel should be prepared.

1. Comparison of protein separation linear range is shown in Schedule 1.
2. Sample volume of each component corresponding to different stacking gel volumes is shown in Schedule 2.
3. Sample volume of each component corresponding to different separation gel volumes is shown in Schedule 3.

**Procedure**

1. Preparation of 10% APS: add distilled water or DI water (add 10µL of water to 1mg APS) to APS [E-IR-R305D] to prepare a 10% APS, For example: add 1.7mL distilled water to 170mg APS, the mixture is 10% APS. 10% APS should be split into small tubes and stored at -20°C.
2. Choose the clean gel mold and complete the assembly.
3. Refer to Schedule 3, prepare the separating gel of the required concentration and volume. Add appropriate amount of ddH<sub>2</sub>O, 30% Acr-Bis (29:1) [E-IR-R305A] and Separating Gel Mix [E-IR-R305B] to a clean beaker, mix it.
4. Add appropriate amount of 10% APS and TEMED, mix gently and avoid bubbles.
5. Add the mixture gel to the assembled gel mold immediately. Then add 1~2 mL ddH<sub>2</sub>O or absolute alcohol on the separating gel to flatten the separating gel level.
6. Keep at RT for 30~60 min until the separating gel solidified completely.
7. Pour out the ddH<sub>2</sub>O or absolute ethanol, absorb the residual liquid with absorbent paper.
8. Refer to Schedule 2, prepare the stacking gel of the required volume. Add appropriate amount of ddH<sub>2</sub>O, 30% Acr-Bis (29:1) [E-IR-R305A] and stacking Gel Mix [E-IR-R305C] to a clean beaker, mix it.
9. Add appropriate amount of 10% APS and TEMED, mix gently and avoid bubbles.
10. Add the mixture gel upon to the separating gel immediately. Insert comb teeth carefully, avoid bubbles.

**For Research Use Only**

11. Keep at RT for 30~40 min until the stacking gel solidified completely.
12. Pull out the comb teeth carefully. Follow the SDS-PAGE experiment.

## Storage

Store separately for 6 months.

30% Acr-Bis (29:1) [E-IR-R305A] and TEMED [E-IR-R305E] should be stored at 2~8°C in dark.

Ammonium persulfate [E-IR-R305D] prepared as a 10% solution, aliquoted, and stored at -20°C has a shelf life of 6 months. If stored at 2~8°C, please do not use it after one week.

Other reagents should be stored at RT.

## Cautions

1. For maximal assay performance, this reagent should be used within 6 months. Avoid freeze / thaw cycles.
2. TEMED is volatile. Please tighten the cap after use. Due to the different temperatures, the rate of gel solidification will be different. The amount of TEMED can be adjusted appropriately to adjust the rate of gelation at different temperatures.
3. Avoid bubbles to avoid affecting the experimental results.
4. Add the ddH<sub>2</sub>O or absolute alcohol on the separating gel gently to avoid uneven line pressing.
5. This kit is for research use only. For your safety and health, please wear lab clothes and gloves. Instructions should be followed strictly, changes of operation may result in unreliable results.

## Schedule

### Schedule 1 Comparison of protein separation linear range

| Concentration of separating gel | Linear separation range |
|---------------------------------|-------------------------|
| 6%                              | 50~150kDa               |
| 8%                              | 30~90kDa                |
| 10%                             | 20~80kDa                |
| 12%                             | 12~60kDa                |
| 15%                             | 10~40kDa                |

## Schedule 2 Preparation of stacking gel

| Components             | Sample volume of each component corresponding to different stacking gel volumes |       |       |       |       |       |       |      |
|------------------------|---|-------|-------|-------|-------|-------|-------|------|
|                        | 1mL   | 2mL   | 3mL   | 4mL   | 5mL   | 6mL   | 8mL   | 10mL |
| <b>5% Stacking Gel</b> |   |       |       |       |       |       |       |      |
| H <sub>2</sub> O       | 0.68  | 1.4   | 2.1   | 2.7   | 3.4   | 4.1   | 5.5   | 6.8  |
| 30% Acr-Bis            | 0.17  | 0.33  | 0.5   | 0.67  | 0.83  | 1.0   | 1.3   | 1.7  |
| Stacking Gel Mix       | 0.14  | 0.27  | 0.41  | 0.54  | 0.68  | 0.81  | 1.08  | 1.35 |
| 10% APS                | 0.01  | 0.02  | 0.03  | 0.04  | 0.05  | 0.06  | 0.08  | 0.1  |
| TEMED                  | 0.001   | 0.002 | 0.003 | 0.004 | 0.005 | 0.006 | 0.008 | 0.01 |

## Schedule 3 Preparation of separating gel

| Components         | Sample volume of each component corresponding to different separation gel volumes |       |       |       |       |       |       |      |
|--------------------|---|-------|-------|-------|-------|-------|-------|------|
|                    | 5mL   | 10mL  | 15mL  | 20mL  | 25mL  | 30mL  | 40mL  | 50mL |
| <b>6% Gel</b>      |   |       |       |       |       |       |       |      |
| ddH <sub>2</sub> O | 2.6   | 5.3   | 7.9   | 10.6  | 13.2  | 15.9  | 21.2  | 26.5 |
| 30% Acr-Bis        | 1.0   | 2.0   | 3.0   | 4.0   | 5.0   | 6.0   | 8.0   | 10.0 |
| Separating Gel Mix | 1.35  | 2.6   | 3.95  | 5.2   | 6.55  | 7.8   | 10.4  | 13.0 |
| 10% APS            | 0.05  | 0.1   | 0.15  | 0.2   | 0.25  | 0.3   | 0.4   | 0.5  |
| TEMED              | 0.004   | 0.008 | 0.012 | 0.016 | 0.02  | 0.024 | 0.032 | 0.04 |
| <b>8% Gel</b>      |   |       |       |       |       |       |       |      |
| ddH <sub>2</sub> O | 2.3   | 4.6   | 6.9   | 9.3   | 11.5  | 13.9  | 18.5  | 23.2 |
| 30% Acr-Bis        | 1.3   | 2.5   | 4.0   | 5.3   | 6.7   | 8.0   | 10.7  | 13.3 |
| Separating Gel Mix | 1.35  | 2.6   | 3.95  | 5.2   | 6.55  | 7.8   | 10.4  | 13.0 |
| 10% APS            | 0.05  | 0.1   | 0.15  | 0.2   | 0.25  | 0.3   | 0.4   | 0.5  |
| TEMED              | 0.003   | 0.006 | 0.009 | 0.012 | 0.015 | 0.018 | 0.024 | 0.03 |
| <b>10% Gel</b>     |   |       |       |       |       |       |       |      |
| ddH <sub>2</sub> O | 1.9   | 4.0   | 5.9   | 7.9   | 9.9   | 11.9  | 15.9  | 19.8 |
| 30% Acr-Bis        | 1.7   | 3.3   | 5.0   | 6.7   | 8.3   | 10.0  | 13.3  | 16.7 |
| Separating Gel Mix | 1.35  | 2.6   | 3.95  | 5.2   | 6.55  | 7.8   | 10.4  | 13.0 |
| 10% APS            | 0.05  | 0.1   | 0.15  | 0.2   | 0.25  | 0.3   | 0.4   | 0.5  |
| TEMED              | 0.002   | 0.004 | 0.006 | 0.008 | 0.01  | 0.012 | 0.016 | 0.02 |
| <b>12% Gel</b>     |   |       |       |       |       |       |       |      |
| ddH <sub>2</sub> O | 1.6   | 3.3   | 4.9   | 6.6   | 8.2   | 9.9   | 15.9  | 16.5 |
| 30% Acr-Bis        | 2.0   | 4.0   | 6.0   | 8.0   | 10.0  | 12.0  | 13.3  | 20.0 |
| Separating Gel Mix | 1.35  | 2.6   | 3.95  | 5.2   | 6.55  | 7.8   | 10.4  | 13.0 |
| 10% APS            | 0.05  | 0.1   | 0.15  | 0.2   | 0.25  | 0.3   | 0.4   | 0.5  |
| TEMED              | 0.002   | 0.004 | 0.006 | 0.008 | 0.01  | 0.012 | 0.016 | 0.02 |
| <b>15% Gel</b>     |   |       |       |       |       |       |       |      |
| ddH <sub>2</sub> O | 1.1   | 2.3   | 3.4   | 4.6   | 5.7   | 6.9   | 9.2   | 11.5 |
| 30% Acr-Bis        | 2.5   | 5.0   | 7.5   | 10.0  | 12.5  | 15.0  | 20.0  | 25.0 |
| Separating Gel Mix | 1.35  | 2.6   | 3.95  | 5.2   | 6.55  | 7.8   | 10.4  | 13.0 |
| 10% APS            | 0.05  | 0.1   | 0.15  | 0.2   | 0.25  | 0.3   | 0.4   | 0.5  |
| TEMED              | 0.002   | 0.004 | 0.006 | 0.008 | 0.01  | 0.012 | 0.016 | 0.02 |