**(FOR RESEARCH USE ONLY. DO NOT USE IT IN CLINICAL DIAGNOSTICS !)**

Catalog No: E-EL-H1094

Product size: 96T/48T/24T/96T*5

# Elabscience® Human VIM (Vimentin) ELISA Kit

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help (info in the header of each page).


Tel: 1-832-243-6086
Fax: 1-832-243-6017
Email: techsupport@elabscience.com
Website: www.elabscience.com


Please refer to specific expiry date from label outside of box.

Please kindly provide us with the lot number (on the outside of the box) of the kit for more efficient service.

Rev V6.0

## Intended Use

Designed for the *in vitro* quantitative determination of Human Vimentin(VIM) concentrations in Serum, EDTA plasma, heparinized plasma and Cell lysate.

## Performance Characteristics

| Parameter | Specification |
|---|---|
| Sensitivity | 4.69 ng/mL |
| Detection Range | 7.81-500 ng/mL |
| Specificity | Specific for Human VIM. No significant cross-reactivity or interference with related analogues was observed |
| Repeatability | Coefficient of variation is < 10% |

## Test Principle

This assay is based on the sandwich enzyme-linked immunosorbent assay (ELISA) technique. Human Vimentin(VIM) present in the sample is specifically bound by an immobilized capture antibody and subsequently recognized by a biotinylated detection antibody, forming a sandwich immune complex. This complex is further associated with an avidin–horseradish peroxidase (HRP) conjugate, enabling enzymatic signal amplification.

Upon reaction with a chromogenic substrate, the HRP enzyme catalyzes the development of a colored product. The resulting signal intensity, measured spectrophotometrically at 450 ± 2 nm, is directly proportional to the concentration of Human VIM in the sample. Quantitative determination is achieved by comparison with a standard curve generated from known concentrations.

## Target Overview

Vimentin (VIM) is a type III intermediate filament protein and constitutes the major cytoskeletal filament of cells of mesenchymal origin, including fibroblasts, endothelial cells, smooth muscle cells and leukocytes. It is a phosphoprotein with an apparent molecular mass of approximately 55-60 kDa and contains a central coiled-coil rod domain flanked by non-helical head and tail regions that assemble into apolar filaments.

Inside the cell, vimentin filaments stabilize cytoplasmic architecture, maintain cell shape and resist mechanical stress, and they position and anchor organelles such as the nucleus and mitochondria. The network is dynamically reorganized by phosphorylation during migration, adhesion and cytokinesis, participating in cell migration, wound closure and cytoskeletal remodeling. Vimentin can also be exposed on the cell surface and released extracellularly; monocyte-derived macrophages secrete it in a process enhanced by tumor necrosis factor alpha and suppressed by interleukin 10, indicating a role in immune regulation.

Because its expression is characteristic of mesenchymal differentiation, vimentin is routinely used in immunohistochemistry to identify tumors of mesenchymal origin and to detect epithelial-mesenchymal transition during tumor progression. Altered expression accompanies wound healing, fibrosis and inflammation, and extracellular vimentin is investigated as a mediator and marker of tissue injury. This ELISA kit enables quantitative determination of vimentin in serum, plasma and other biological samples.

## Kit Components & Storage

An unopened kit should be stored at 2-8 °C for 12 months. After opening, components should be stored under the conditions specified below:

| Item | Specifications | Storage |
|---|---|---|
| Micro ELISA Plate (Dismountable) | 96T: 8 × 12 strips<br>48T: 8 × 6 strips<br>24T: 8 × 3 strips<br>96T*5: 96T × 5 plates | -20 °C, stable until expiry (12 months) |
| Reference Standard | 96T: 2 vials<br>48T/24T: 1 vial<br>96T*5: 10 vials | |
| Concentrated Biotinylated Detection Ab(100×) | 96T: 120 µL × 1 vial<br>48T/24T: 60 µL × 1 vial<br>96T*5: 120 µL × 5 vials | |
| Concentrated HRP Conjugate (100×) | 96T: 120 µL × 1 vial<br>48T/24T: 60 µL × 1 vial<br>96T*5: 120 µL × 5 vials | |
| Reference Standard & Sample Diluent | 96T/48T/24T: 20 mL × 2 vials<br>96T*5: 20 mL × 10 vials | 2-8 °C, stable until expiry (12 months) |
| Biotinylated Detection Ab Diluent | 96T/48T/24T: 14 mL × 1 vial<br>96T*5: 14 mL × 5 vials | |
| HRP Conjugate Diluent | 96T/48T/24T: 14 mL × 1 vial<br>96T*5: 14 mL × 5 vials | |
| Concentrated Wash Buffer(25×) | 96T/48T/24T: 30 mL × 1 vial<br>96T*5: 30 mL × 5 vials | |
| Substrate Reagent | 96T/48T/24T: 10 mL × 1 vial<br>96T*5: 10 mL × 5 vials | |
| Stop Solution | 96T/48T/24T: 10 mL × 1 vial<br>96T*5: 10 mL × 5 vials | |
| Plate Sealing Film | 96T/48T/24T: 5 pieces<br>96T*5: 25 pieces | |
| Manual | 1 copy | |

**Note**

Reagents may contain slightly more volume than indicated on the label; mix and centrifuge the vial of reagents before use. Use precise measuring instruments rather than pouring directly from the vial.

## Other Supplies Required

- Microplate reader capable of measuring absorbance at 450 nm
- High-precision pipettes, EP tubes, and disposable pipette tips
- Incubator capable of maintaining 37 ℃
- Deionized or distilled water
- Absorbent paper
- Loading slot (Suction Tank)

## Precautions

1. **For Research Use Only**
   - This kit is intended for research purposes and is not for diagnostic or therapeutic use.
2. **Laboratory Safety**
   - Always wear lab coats, safety goggles, and disposable gloves when handling reagents or biological samples.
   - Follow local and national biosafety regulations, especially when working with blood, plasma, or other bodily fluids.
3. **Reagent Handling**
   - Protect light-sensitive reagents such as **Concentrated HRP Conjugate (100×)** and **Substrate Reagent** from exposure to light.
   - Ensure all reagent caps are tightly closed to prevent evaporation and microbial contamination.
   - Do **not reuse** reconstituted standards, working solutions of detection antibody, or HRP conjugate. Stock solutions must be stored according to specified conditions.

- Do not mix reagents from different kit lots or sources.

4. **Pipetting and Cross-Contamination Prevention**
   - Use fresh tips for each standard, sample, and reagent addition.
   - Use separate reservoirs for each reagent to prevent cross-contamination.
5. **Kit Expiry**
   - Do not use the kit beyond the expiration date indicated on the label.

## Sample Collection & Handling

Proper sample handling and collection are critical for the accurate measurement of Human VIM. The following guidelines are recommended:

**General Guidelines**

- Use disposable, endotoxin-free tubes for blood collection.
- Avoid samples with high hemolysis or excessive lipids, as they may interfere with the assay.
- Predict sample concentration beforehand to determine appropriate dilutions.

**Storage and Stability**

- For longer storage: -20 °C for up to 1 month, −80 °C for up to 3 months.
- Avoid repeated freeze-thaw cycles. Thaw frozen samples slowly and centrifuge to remove precipitates before assay.

**Sample Preparation Considerations**

- When using lysis buffers for tissue homogenates or cell lysates, note that chemical components may affect assay results.
- If the sample type is not listed in the manual, perform preliminary validation experiments.
- Some recombinant proteins may not be detected due to mismatches with coated or detection antibodies.

## Sample Types and Preparation Guidelines

**Serum**

- Collect blood in pyrogen-free, endotoxin-free coagulation tubes or plain tubes.
- Allow clotting, at room temperature, for 2 hours, or overnight, at 2-8 °C.
- Centrifuge at 1,000 × g for 20 minutes. For hyperlipidemic samples, centrifuge at 10,000 × g at 2-8 °C for 10 minutes to separate serum and remove chylomicrons.
- Aliquot the supernatant (serum) into sterile cryopreservation tubes.

**Plasma**

- Use EDTA-K2/Na2 (1.5-2 mg/mL) as anticoagulant; sodium citrate can be used for coagulation-related targets.
- Gently invert tubes 5-8 times immediately after collection to prevent air bubbles and hemolysis.
- Centrifuge at 1,000 × g at 2-8 °C for 15 minutes, within 30 minutes after collection.

- Collect the plasma fraction after centrifugation.

**Tissue Homogenate**

- Rinse tissue 2-3 times with pre-cooled PBS (0.01 M, pH 7.2-7.4) to remove residual blood.
- Cut tissue into small pieces (2-3 $mm^3$) and add to a glass homogenizer with PBS containing protease inhibitor (add phosphatase inhibitor for phosphorylated targets; tissue-to-homogenate ratio 1:9).
- Homogenize on ice 30-50 times. For further lysis, repeat freeze-thaw cycles or use ultrasonic disruption.
- Centrifuge at 5,000 × g at 2–8 °C for 5–10 minutes and collect supernatant.
- Optional: Determine total protein concentration using a BCA Protein Quantification Kit.

**Cell Lysate**

- Gently rinse adherent cells with cold PBS; digest with trypsin if needed.
- Collect cells by centrifugation at 1,000 × g for 5 minutes; suspension cells can be collected directly.
- Wash cells 3 times with cold PBS and resuspend in 150-200 µL PBS per $1.0 \times 10^6$ cells (add protease inhibitor if required).
- Lyse cells using repeated freeze-thaw cycles or ultrasonic disruption.
- Centrifuge at 1,500 × g at 2-8 °C for 10 minutes and collect supernatant.

**Cell Culture Supernatant**

- Collect aseptically and centrifuge at 1,000 × g at 2-8 °C for 20 minutes to remove debris.
- Collect the clarified supernatant for ELISA measurement.

**Urine**

- Centrifuge at 1,500 × g at 2-8 °C for 15 minutes to remove particulate matter.
- Use supernatant immediately or aliquot and store at ≤ -20 °C.

**Recommended Reagents for Sample Processing**

- PMSF Protease Inhibitor (Cat No. E-EL-SR002)
- 0.25% Trypsin Solution (Cat No. E-EL-SR001)
- BCA Protein Quantification Kit (Cat No. E-BC-K318-M)

## Sample Validation Data

**This kit has been validated using samples from apparently healthy volunteers.**

**The data is for reference only.**

| Species | Sample Type | Clinical Characteristics & Group | Recommended Dilution Factor | Reference Range |
|---|---|---|---|---|
| Human | Serum (n=30) | Healthy individuals | Undiluted | 9.22-243.25 ng/mL |
| Human | EDTA plasma (n=30) | Healthy individuals | Undiluted | 17.72-56.51 ng/mL |

## Sample Dilution Protocol for Dilution Factors > 100-Fold

The following is the general procedure that can be adjusted based on sample volume and dilution factor:

**General Principles**

- Serial dilution is recommended; the dilution factor at each step should not exceed 100-fold.

  Pipetting volume at each step should be ≥ 3 μL. Mix thoroughly and avoid air bubbles.

**Procedure Examples**

**100-fold dilution:** One-step dilution. Add 5 μL sample to 495 μL sample diluent.

**1,000-fold dilution:** Two-step dilution. Add 5 μL sample to 95 μL sample (20-fold), then 5 μL 20-fold diluted sample to 245 μL sample diluent (final 1,000-fold).

**10,000-fold dilution:** Two-step dilution. Add 5 μL sample to 495 μL sample diluent (100-fold), then 5 μL 100-fold diluted sample to 495 μL sample diluent (final 10,000-fold).

**100,000-fold dilution:** Three-step dilution. Add 5 μL sample to 195 μL sample diluent (40-fold), then 5 μL 40-fold diluted sample to 245 μL sample diluent (50-fold dilution), and finally 5 μL 2,000-fold diluted sample to 245 μL sample diluent (final: 100,000-fold).

Higher dilution factors can be achieved by increasing the number of serial dilutions, following the same principles. Select the dilution strategy based on the sample volume and estimated concentration to ensure accuracy and repeatability.

## Reagent preparation

**1. General Handling**

- Equilibrate all reagents to room temperature (18-25°C) before use.
- If the kit is to be used partially in a single assay, take only the strips and reagents required for the current assay, leaving the remainder under the conditions specified in the storage table.


2. **Wash Buffer**

- Prepare the working Wash Buffer by diluting 30 mL of Concentrated Wash Buffer with 720 mL of deionized or distilled water yielding 750 mL.
- Note: If crystalline precipitates have formed in the concentrate, warm it gently in a 40 °C water bath and mix until fully dissolved. Use the prepared Wash Buffer on the same day only.

3. **Standard Working Solution**

- Centrifuge the standard at 10,000 × g for 1 minute.
- Reconstitute with 1 mL of the Standard & Sample Diluent, allow to stand for 10 minutes, and gently invert several times until completely dissolved. Alternatively, vortex briefly at low speed; remove bubbles by low-speed centrifugation if necessary.
- This produces a 500 ng/mL stock solution. Prepare serial dilutions according to experimental requirements. Recommended dilution gradient: 500, 250, 125, 62.5, 31.25, 15.63, 7.81, 0 ng/mL.

- Serial Dilution Procedure:
    i. Dispense 500 µL of Standard & Sample Diluent into 7 EP tubes, respectively.
    ii. Transfer 500 µL of the 500 ng/mL stock solution into the first tube to obtain 250 ng/mL.
    iii. Sequentially transfer 500 µL from the preceding tube into the next tube for subsequent dilutions. The illustration on the next page is provided for reference. Designate the last tube as the blank; do not transfer any solution into it.
- Aliquot the 500 ng/mL stock and store at -20 °C. Use within two weeks and avoid repeated freeze-thaw cycles.
- Serially diluted standards should be prepared immediately, prior to use.



4. **Biotinylated Detection Ab Working Solution**

- Calculate the required volume (100 µL/well) plus a slight excess.
- Centrifuge the Concentrated Biotinylated Detection Antibody (100×) at 800 × g for 1 minute.
- Dilute to 1× with Biotinylated Detection Ab Diluent at a ratio of 1:99.
- Prepare the working solution immediately, prior to use.

5. **HRP Conjugate Working solution:**

- Calculate the required volume (100 µL/well) plus a slight excess.
- Centrifuge the Concentrated HRP Conjugate (100×) at 800 × g for 1 minute.
- Dilute to 1× with HRP Conjugate Diluent at a ratio of 1:99.
- Prepare the working solution immediately, prior to use.

## Assay procedure

1. **Sample and Standard Loading**

- Assign wells for diluted standards, blanks, and samples.
- Add 100 µL of each standard, blank, or sample to the designated wells. Analyze all samples and standards in duplicate. Determine sample dilution factors based on preliminary experiments or technical support recommendations.
- Cover the plate with the supplied sealer and incubate for 90 minutes at 37 °C.
- Note: Dispense solutions gently to the bottom of the well, avoiding contact with the sidewalls and foaming.

2. **Addition of Biotinylated Detection Antibody**

- Remove liquid from each well without washing.
- Add 100 µL of Biotinylated Detection Ab Working Solution to each well. Cover with a fresh sealer and incubate for 60 minutes at 37 °C.

3. **Washing Step 1**

- Decant the solution and add 350 µL of Wash Buffer to each well. Soak for 1 minute, then aspirate or decant and blot dry with clean absorbent paper.
- Repeat three times.
- Note: A microplate washer may be used. Do not allow wells to dry. Proceed immediately with the next step.

4. **Addition of HRP Conjugate**

- Add 100 µL of **HRP Conjugate Working Solution** to each well. Cover the plate with a new sealer. Incubate for 30 minutes at 37 °C.

5. **Washing Step 2**

- Repeat the wash procedure, described in Step 3, five times.

6. **Substrate Reaction**

- Add 90 µL of **Substrate Reagent** to each well. Cover and incubate for approximately 15 minutes at 37°C. Protect the plate from light.
- Note: Reaction time may be adjusted based on observed color development but should not exceed 30 minutes. Switch on the

Microplate Reader for 15 minutes prior to OD measurement.

7. **Stopping the Reaction**

- Add 50 µL of **Stop Solution** to each well in the same order as the substrate addition.

8. **Optical Density Measurement**

- Measure the optical density (OD value) at 450 ± 2 nm using a microplate reader immediately after adding Stop Solution .






**Figure: Example Plate Layout**

**ST: Standard curve well; Blank: Blank well (0 ng/mL).**

**S: Sample well.**

# Assay Procedure Summary




1. Add 100μL standard or sample to the wells. Incubate for 90 min at 37°C

2. Discard the liquid, immediately add 100μL Biotinylated Detection Ab working solution to each well. Incubate for 60 min at 37°C

3. Aspirate and wash the plate for 3 times

4. Add 100μL HRP conjugate working solution. Incubate for 30 min at 37°C. Aspirate and wash the plate for 5 times

5. Add 90μL Substrate Reagent. Incubate for 15 min at 37°C

6. Add 50μL Stop Solution

7. Read the plate at 450nm immediately. Calculation of the results

## Calculation of results

1. **Data Processing**
   - Calculate the mean optical density (OD) of duplicate readings for each standard and sample.
   - Subtract the mean OD of the zero standard (blank) from all measurements.

2. **Standard Curve**
   - Plot a four-parameter logistic (4PL) curve with standard concentrations on the x-axis (log scale optional) and OD values on the y-axis.


3. **Sample Concentration**
   - Determine the concentration of each sample from the standard curve.
   - If the sample OD exceeds the upper limit of the standard curve, repeat the assay using an appropriate dilution.
   - The final concentration is calculated using the following equation:

   *Final concentration = Measured concentration × Dilution factor*

## Typical data

The following data was generated by the Quality Control Department, under controlled laboratory conditions (ambient temperature: 18-25 °C, relative humidity: 35-75%) using standardized procedures (TMB reaction at 37 °C in the dark for 15 minutes, followed by termination and OD measurement). These values are provided for reference only. Actual results may vary due to differences in laboratory conditions, operator technique, and equipment. Users are required to generate a standard curve using their own experimental data.

| ng/mL | OD1 | OD2 | Mean OD | Corrected OD |
|---|---|---|---|---|
| 500 | 2.553 | 2.525 | 2.539 | 2.447 |
| 250 | 1.672 | 1.694 | 1.683 | 1.591 |
| 125 | 1.003 | 0.993 | 0.998 | 0.906 |
| 62.5 | 0.529 | 0.505 | 0.517 | 0.425 |
| 31.25 | 0.334 | 0.330 | 0.332 | 0.240 |
| 15.63 | 0.198 | 0.204 | 0.201 | 0.109 |
| 7.81 | 0.146 | 0.150 | 0.148 | 0.056 |
| 0 | 0.094 | 0.090 | 0.092 | - |



## Performance

### ■ Precision

Intra-assay Precision (Within-run Precision): Three samples representing low, mid, and high concentrations of Human VIM were tested 20 times on a single plate.

Inter-assay Precision (Between-run Precision): Three samples representing low, mid, and high concentrations of Human VIM were tested on three separate plates, with 20 replicates per plate, to assess variability among assays.

| | Intra-assay Precision | | | Inter-assay Precision | | |
|---|---|---|---|---|---|---|
| Sample | 1 | 2 | 3 | 1 | 2 | 3 |
| Numbers | 20 | 20 | 20 | 20 | 20 | 20 |
| Mean (ng/mL) | 27.05 | 71.67 | 178.89 | 25.06 | 73.92 | 192.36 |
| Standard deviation | 1.58 | 3.48 | 10.75 | 1.62 | 4.83 | 15.6 |
| CV (%) | 5.84 | 4.85 | 6.01 | 6.46 | 6.53 | 8.11 |

■ **Recovery**

The recovery of Human VIM was evaluated by spiking samples at low, mid, and high concentrations across the assay range in various sample matrices. The assay performance was assessed by comparing the measured concentrations to the expected spiked amounts to determine the percent recovery.

| Sample Type | Range (%) | Average Recovery (%) |
|---|---|---|
| Serum (n=8) | 95-109 | 101 |
| EDTA plasma (n=8) | 89-101 | 94 |
| Cell culture media (n=8) | 89-101 | 95 |

■ **Linearity**

Linearity of the assay was evaluated by spiking samples with high concentrations of Human VIM and performing serial dilutions using Standard & Sample Diluent to produce concentrations spanning the assay's dynamic range. The measured values were then compared to the expected concentrations to assess the linearity of response.

| | | Serum (n=5) | EDTA plasma (n=5) | Cell culture media(n=5) |
|---|---|---|---|---|
| 1:2 | Range (%) | 86-101 | 87-99 | 94-106 |
| | Average (%) | 92 | 94 | 100 |
| 1:4 | Range (%) | 89-103 | 82-93 | 83-94 |
| | Average (%) | 96 | 87 | 88 |
| 1:8 | Range (%) | 88-101 | 79-89 | 87-99 |
| | Average (%) | 93 | 85 | 94 |
| 1:16 | Range (%) | 94-108 | 83-96 | 83-93 |
| | Average (%) | 100 | 89 | 88 |

■ **Stability**

Each kit batch is subjected to accelerated stability testing and real-time stability monitoring. Sample performance is evaluated after storage at 37 °C for 10 days to assess the impact of elevated temperature on assay reliability and reagent integrity.

| | Variation range of 37°C mean concentration / 2-8°C mean concentration (%) |
|---|---|
| Sample 1 (n=16) | 92.21-107.34 |
| Sample 2 (n=16) | 92.52-103.35 |

## Declaration

1. Due to current technological and methodological limitations, comprehensive identification and analysis of all raw materials cannot be guaranteed. Users should be aware of potential qualitative and technical risks associated with kit use.
2. This assay is designed to minimize interference by factors present in biological samples; however, until all potential interfering substances are fully evaluated, their influence cannot be entirely excluded.
3. Experimental results are dependent on reagent quality, operator technique, and laboratory conditions. The manufacturer is responsible only for the performance of the kit itself, not for the samples utilized during the assay. Users should estimate and reserve sufficient sample volume for their experiments.
4. Optimal results are achieved only when all reagents supplied with the kit are used and the instructions are strictly followed.
5. Inaccurate results may occur due to improper reagent preparation, incorrect sample or reagent loading, or misconfigured microplate reader parameters. Users should carefully review instructions and calibrate instruments prior to assay.
6. Variability may occur even when the same operator performs separate assays. Consistent execution of all procedural steps is essential for reproducible results.
7. All kits undergo rigorous quality control; however, variations in results may arise from differences in transportation, laboratory equipment, or environmental conditions. Intra-assay variability may also occur between different kit batches.

8. Results may differ when using kits from other manufacturers or alternative assay methods for the same analyte, as comparative evaluations have not been performed.
9. This kit is intended for research use only. The manufacturer is not responsible for any outcomes resulting from use in clinical diagnostics or other non-research applications.