

Duck/Goose-Derived Material Lateral Flow Assay Kit

Catalog No: E-FS-C138

20T/40T/80T

| | |
|-------------------------|------------|
| Version Number: | V1.0 |
| Replace version: | / |
| Revision Date: | 2025.06.17 |

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Toll-free: 1-888-852-8623 Tel: 1-832-243-6086 Fax: 1-832-243-6017

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Test principle

This kit uses the principle of sandwich immunochromatography assay for the qualitative detection. It can detect of duck/goose source components in the samples.

Kits components

| Item | Specifications |
|--------------------------|----------------|
| Detection Card | 40 T/kit |
| Sample Extraction Buffer | 40 vials |
| Swab | 40 pieces |
| Dropper | 40 pieces |
| Manual | 1 copy |

Note: All reagent bottle caps must be tightened to prevent evaporation and microbial pollution.

Other supplies required

Instruments: Homogenizer, Centrifuge, Graduated pipette, Balance (sensibility 0.01g), Oscillator, Water bath.

High-precision transferpettor: Single channel (20-200 µL, 100-1000 µL).

Notes

1. FOR RESEARCH USE ONLY. Do not use product out of date or in a broken aluminum foil.
2. Bring detection card to room temperature before opening the aluminum foil. The opening detection card should be used as soon as possible so as not to be invalid because of moisture.
3. Avoid of contacting the white membrane at the middle of the sample well.
4. The disposable dropper cannot be mixing to avoid the cross-contaminant.
5. The tested sample should be clear, no turbidity particle and no bacterial pollution, otherwise it is easy to result in abnormal phenomena such as obstruction and unobvious color which affect the judgment of the experiment result.
6. If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.
7. **Each reagent is optimized for use in the E-FS-C138. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other E-FS-C138 with different lot numbers.**
8. The kit is used for rapid screening of actual samples. If the test result is positive, the instrument method such as HPLC, LC/MS, etc. can be used for quantitative confirmation.

Storage and expiry date

Storage: Store at 2-30°C. With cool and dry environment.

Expiry date: expiration date is on the packing box.

Sample pretreatment

Restore all reagents and samples to room temperature before use.

1. Sample pretreatment Notice:

Experimental apparatus should be clean, and the disposable dropper should be disposable to avoid the experiment result be interfered by the contamination.

2. Pretreatment of the surface tissue fluid of meat

- (1) Using a swab, absorb tissue fluid from the surface of the sample to be tested. Then soak the swab in the extraction solution for 10 s. Vigorously stir and mix the solution for 10-20 s (moving the swab up, down, left, and right) to ensure the sample dissolves as completely as possible.
- (2) Remove the swab.

3. Pretreatment of muscle

- (1) Use scissors to cut 0.1 ± 0.05 g (about the size of a soybean) of meat pieces. Add the meat pieces to the extraction solution and let them soak for 10 s. Then, use a swab to squeeze the meat pieces 5 to 6 times and stir thoroughly up, down, left and right for 10 to 20 s. This is the solution to be tested.

Experiment procedure

1. Tear the aluminum foil bag of the detection card and take out the detection card, dropper and put it on a smooth, clean table.
2. Take 100 μ L (about 3 droppers) sample into the sample well of detection card.

3. Allow it to react at room temperature for 5-10 min before interpreting the results, the result out of 15 min is invalid.

Judgment of result

1. **Negative:** The control line region (C) show color, the test line region duck source (Y) line or/and the goose source (E) line shows no color, and the test result is negative.
2. **Positive:** The control line region (C) show color, the test line region duck source (Y) line or/and the goose source (E) line shows color, and the test result is positive.
3. **Invalid:** The control line region (C) show no color. It indicates operation process is wrong or the test card is invalid.

