

SAs (Sulfonamides) Lateral Flow Assay Kit

Catalog No: E-FS-C130

40T

Version Number:	V1.3
Replace version:	V1.2
Revision Date:	2026.04.28

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Toll-free: 1-888-852-8623 Tel: 1-832-243-6086 Fax: 1-832-243-6017

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Test principle

This kit uses the principle of Immunochromatography assay for the qualitative detection. It can detect SAs (Sulfonamides) in raw milk, milk, milk powder. After adding the sample solution into the sample well of detection card, SAs of the sample solution combine with the gold-labelled antibody, so as to prevent the combining of gold-labelled antibody with SAs conjugate on the cellulose membrane. When the concentration of SAs in the sample solution is more than the detection limit, the detect line do not show color (or shows lighter color than control line) and the result is positive. When the concentration of SAs in the sample solution is less than the detection limit, the detect line shows color (the color is equal or darker than the control line) and the result is negative.

Technical indicator

Detection limit of sulfonamides:

Name	Detection limit (ppb)
Sulfachlorpyridazine	2-4
Sulfadiazine	2.5-5
Sulfisomidine / Sulfadimetine	1-2
Sulfadimethoxine / Sulfadimethoxin / Sulfamethoxine / Sulfadimethoxypyrimidine	2.5-5
Sulfamethoxypyridazine	10-20
Sulfamerazine	2-4
Sulfamonomethoxine	1-2
Sulfadimidine / Sulfamethazine	1-2
Sulfamethizole	6-12
Sulfametoxydiazine	1-2
Sulfaquinoxaline	3-6
Sulfachloropyrazine	2.5-5
Sulfadimethyloxazole / Sulfoxazole	8-15

Kits components

Item	Specifications
The Test Strip	40 T/kit
Gold-labelled micro well	40 wells
Manual	1 copy

Other materials required but not supplied

Instruments: Homogenizer, Oscillators, Centrifuge, Graduated pipette, Incubator, Balance (sensitivity 0.01g)

High-precision transferpettor: Single channel (20-200 μ L, 100-1000 μ L)

Notes

1. FOR RESEARCH USE ONLY. Do not use product out of date or in a broken aluminum foil.
2. The detection card should be adjusted to room temperature after removed from the refrigerator before opening. The opening detection card should be used as soon as possible so as not to be invalid because of moisture.
3. Avoid of contacting the white membrane at the middle of the sample well.
4. The pipette cannot be mixing to avoid the cross-contaminant.
5. The tested sample should be clear, no turbidity particle and no bacterial pollution, otherwise it is easy to result in abnormal phenomena such as obstruction, unobvious color, etc., which affect the judgment of the experiment result.
6. If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.
7. The kit is used for rapid screening of actual samples. If the test result is positive, the instrument method such as HPLC, LC/MS, etc. can be used for quantitative confirmation.
8. Each reagent is optimized for use in the E-FS-C130. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other E-FS-C130 with different lot numbers.

Storage and expiry date

Storage: Store at 2-30°C. With cool and dry environment.

Expiry date: expiration date is on the packing box.

Sample pretreatment

Restore all reagents and samples to room temperature before use.

1. Sample pretreatment Notice:

Experimental apparatus should be clean, and the pipette should be disposable to avoid the experiment result be interfered by the contamination.

2. Sample pretreatment procedure:

2.1 Pretreatment of milk powder sample:

Weigh 1.00 ± 0.05 g of milk powder into a 10 mL centrifuge tube, add 8 mL of deionized water, and mix well.

2.2 Pretreatment of raw milk and milk sample:

The raw milk and milk sample must not have lumps, fermentation, sedimentation or other properties.

Experiment procedure

1. Tear the aluminum foil bag of the detection card and take out the detection card, and put it on a smooth, clean table.
2. Shake the milk sample to be tested thoroughly. Take 200 μL into the gold-labelled micro well. Carefully pipette until the purple-red particles at the bottom of the well are completely dissolved (Note: Pipetting should be gentle to avoid air bubbles). Then place the gold-labelled micro well in a 50°C incubator and incubate for 3 minutes.
3. Insert the test strip into the gold-labeled microwell to fully immerse the sample pad in the milk sample, and allow the reaction to proceed for 5 minutes.
4. Take out the test strip from the gold-labeled microwell, remove the sample pad at the bottom of the test strip, and interpret the test results.

Judgment of result

1. **Negative:** The control line region (C) show color, the test line region (T) shows equal or darker than line C. It indicates the content of SAs in the sample is lower than detection limit or the sample doesn't contain SAs.
2. **Positive:** The control line region (C) show color, the test line region (T) shows no color or lighter color than line C. It indicates the content of SAs in the sample is higher than detection limit.
3. **Invalid:** The control line region (C) shows no color. It indicates operation process is wrong or the test card is invalid.

