

(FOR RESEARCH USE ONLY. DO NOT USE IT IN CLINICAL DIAGNOSIS !)

Catalog No: E-BC-K007-M

Specification: 48T(16 samples) / 96T(40 samples)/ 500Assays(242 samples)

Measuring instrument: Microplate reader (535-540 nm)

Detection range: 0.0097-0.3474 U/mL

Elabscience[®] α -Amylase Activity Assay Kit

This manual must be read attentively and completely before using this product.

If you have any problem, please contact our Technical Service Center for help:

Toll-free: 1-888-852-8623

Tel: 1-832-243-6086

Fax: 1-832-243-6017

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Table of contents

| | |
|---|-----------|
| Assay summary | 3 |
| Intended use | 4 |
| Detection principle | 4 |
| Kit components & storage | 4 |
| Materials prepared by users | 4 |
| Reagent preparation | 5 |
| Sample preparation | 6 |
| The key points of the assay | 7 |
| Operating steps | 7 |
| Calculation | 8 |
| Appendix I Performance Characteristics | 10 |
| Appendix II Example Analysis | 12 |
| Statement | 13 |

Assay summary



Intended use

This kit can measure α -amylase activity in serum, saliva, animal and plant tissue samples.

Detection principle

The reducing sugar reacts with 3,5-dinitrosalicylic acid under heating conditions to produce a brown-red substance, which is inactivated by the thermolabile nature of β -amylase, and then the enzyme activity of α -amylase is determined.

Kit components & storage

| Item | Component | Size 1 (48 T) | Size 2 (96 T) | Size 3 (500 Assays) | Storage |
|-----------|-------------------|------------------|------------------|------------------------|---------------------------------------|
| Reagent 1 | Substrate | 5 mL×1 vial | 10 mL×1 vial | 50 mL×1 vial | 2-8°C, 12 months |
| Reagent 2 | Chromogenic Agent | 10 mL×1 vial | 20 mL×1 vial | 50 mL×2 vials | 2-8°C, 12 months, shading light |
| Reagent 3 | 10 mg/mL Standard | 1.5 mL×1 vial | 1.5 mL×1 vial | 7.5 mL×1 vial | 2-8°C, 12 months |
| | Microplate | 48 wells | 96 wells | / | No requirement |
| | Plate Sealer | 2 pieces | | | |

Note: The reagents must be stored strictly according to the preservation conditions in the above table. The reagents in different kits cannot be mixed with each other. For a small volume of reagents, please centrifuge before use, so as not to obtain sufficient amount of reagents.

Materials prepared by users

Instruments:

Test tubes, Vortex Mixer, Centrifuge, Water bath, Microplate reader (535-540 nm, optimum wavelength: 540 nm)

Reagent preparation

- ① Equilibrate all the reagents to room temperature before use.
- ② If there is precipitation in substrate and chromogenic agent, Heat it at 70-80°C in water bath until dissolved. Cool down to 40°C with fresh water before use.
- ③ The preparation of standard curve:

Always prepare a fresh set of standards. Discard working standard dilutions after use.

Dilute 10 mg/mL standard solution with double distilled water to a serial concentration. The recommended dilution gradient is as follows: 0, 0.2, 0.4, 0.6, 0.8, 1.0, 1.2, 1.4 mg/mL. Reference is as follows:

| Item | ① | ② | ③ | ④ | ⑤ | ⑥ | ⑦ | ⑧ |
|------------------------------------|----------|------------|------------|------------|------------|------------|------------|------------|
| Concentration (mg/mL) | 0 | 0.2 | 0.4 | 0.6 | 0.8 | 1.0 | 1.2 | 1.4 |
| 10 mg/mL standard (μL) | 0 | 20 | 40 | 60 | 80 | 100 | 120 | 140 |
| Double distilled water (μL) | 1000 | 980 | 960 | 940 | 920 | 900 | 880 | 860 |

Sample preparation

① Sample preparation

Serum and plasma: detect directly. If not detected on the same day, the serum or plasma can be stored at -80°C for a month.

Saliva sample: Gargle with clear water, collect the saliva 30 min later, centrifuge at 10000×g for 10 min at 4°C. Take the supernatant and preserve it on ice for detection.

Tissue sample:

- ① Weigh 0.1 g sample, add 0.9 mL of distilled water and homogenized with a homogenizer in ice water bath, then transfer to the EP tube, incubate at room temperature for 15 min and oscillate every 5 min.
- ② Centrifuge at 3000×g at room temperature for 10 min, take the supernatant and add double distilled water to the final volume of 10 mL and it is the prepared sample.

② Dilution of sample

The recommended dilution factor for different samples is as follows (for reference only):

| Sample type | Dilution factor |
|---|-----------------|
| 1% <i>Epipremnum aureum</i> tissue homogenate | 1 |
| 1% Green pepper tissue homogenate | 1 |
| 1% Corn grain tissue homogenate | 1 |
| 1% <i>Daucus carota</i> tissue homogenate | 1 |

Note: The diluent is double distilled water. For the dilution of other sample types, please do pretest to confirm the dilution factor.

The key points of the assay

- ① For measuring the OD value, if there is precipitation, centrifuge at $4000\times g$ for 5 min at room temperature and take the supernatant for determination.
- ② When the absolute OD value is more than 0.747, it is recommended to dilute the sample.

Operating steps

1. The measurement of standard

- ① Take 1.5 mL EP tube and number the tubes from A to H in duplication, add 75 μL of standard solution with different concentrations to the corresponding tubes.
- ② Add 75 μL of substrate to each tube.
- ③ Add 150 μL of chromogenic agent to each tube.
- ④ Mix fully and incubate at 95°C for 5 min. Cool the tubes with running water and take 250 μL of supernatant to the microplate. Measure the OD value of each well with microplate reader at 540 nm.

2. The measurement of sample

- ① Sample tube: add 75 μL of sample to the corresponding tubes.
Control tube: add 75 μL of sample to the corresponding tubes.
- ② Incubate at 70°C water bath for 15 min and cool the tubes with running water.
- ③ Sample tube: add 75 μL of substrate to the corresponding tubes.
Control tube: add 75 μL of double distilled water to the corresponding tubes.
- ④ Incubate the sample tubes and control tubes at 40°C water bath for 5 min.
- ⑤ Add 150 μL of chromogenic agent to each tube.
- ⑥ Mix fully and incubate at 95°C for 5 min. Cool the tubes with running water and take 250 μL of supernatant to the microplate. Measure the OD value of each well with microplate reader at 540 nm.

Calculation

The standard curve:

1. Average the duplicate reading for each standard.
2. Subtract the mean OD value of the blank (Standard #①) from all standard readings. This is the absolved OD value.
3. Plot the standard curve by using absolved OD value of standard and correspondent concentration as y-axis and x-axis respectively. Create the standard curve ($y = ax + b$) with graph software (or EXCEL).

The sample:

1. Serum (plasma) and other saliva sample:

Definition: The production of 1 mg reducing sugar catalyzed by 1 mL of serum, saliva per minute that is defined as an enzyme activity unit.

$$\alpha\text{-amylase activity (U/mL)} = (\Delta A - b) \div a \times V_3 \div t \div V_2 \times f$$

2. Tissue sample:

- 1) Calculate according to the protein concentration of the sample

Definition: The production of 1 mg reducing sugar catalyzed by 1 mg of tissue protein per minute that is defined as an enzyme activity unit.

$$\alpha\text{-amylase activity (U/mgprot)} = (\Delta A - b) \div a \times V_3 \div t \div V_2 \times f \div C_{pr}$$

- 2) Calculate according to the fresh weight of sample

Definition: The production of 1 mg reducing sugar catalyzed by 1 g of tissue per minute that is defined as an enzyme activity unit.

$$\alpha\text{-amylase activity (U/g tissue)} = (\Delta A - b) \div a \times V_3 \div t \div w \times \frac{V_1}{V_2} \times f$$

[Note]

f: Dilution factor of sample before tested.

ΔA : $OD_{\text{Sample}} - OD_{\text{Control}}$.

V_1 : The volume of prepared tissue sample in sample preparation step (10 mL).

V_2 : The volume of sample added to the reaction (0.075 mL).

V_3 : The volume of enzymatic reaction (the volume of sample + the volume of substrate = 0.15 mL).

t: The time of enzymatic reaction (5 min).

w: The weight of tissue sample (0.1 g).

C_{pr} : Concentration of protein in sample (mgprot/mL).

Appendix I Performance Characteristics

1. Parameter:

Intra-assay Precision

Three human serum samples were assayed in replicates of 20 to determine precision within an assay. (CV = Coefficient of Variation)

| Parameters | Sample 1 | Sample 2 | Sample 3 |
|-------------|----------|----------|----------|
| Mean (U/mL) | 0.02 | 0.15 | 0.30 |
| %CV | 3.3 | 2.8 | 2.9 |

Inter-assay Precision

Three human serum samples were assayed 20 times in duplicate by three operators to determine precision between assays.

| Parameters | Sample 1 | Sample 2 | Sample 3 |
|-------------|----------|----------|----------|
| Mean (U/mL) | 0.02 | 0.15 | 0.30 |
| %CV | 4.1 | 4.0 | 3.6 |

Recovery

Take three samples of high concentration, middle concentration and low concentration to test the samples of each concentration for 6 times parallelly to get the average recovery rate of 98%.

| | Standard 1 | Standard 2 | Standard 3 |
|------------------------|------------|------------|------------|
| Expected Conc. (mg/mL) | 0.25 | 0.7 | 1.1 |
| Observed Conc. (mg/mL) | 0.3 | 0.7 | 1.1 |
| Recovery rate (%) | 100 | 95 | 99 |


Sensitivity

The analytical sensitivity of the assay is 0.0097 U/mL. This was determined by adding two standard deviations to the mean O.D. obtained when the zero standard was assayed 20 times, and calculating the corresponding concentration.

2. Standard curve:

As the OD value of the standard curve may vary according to the conditions of the actual assay performance (e.g. operator, pipetting technique or temperature effects), so the standard curve and data are provided as below for reference only:

| Concentration (mg/mL) | 0 | 0.2 | 0.4 | 0.6 | 0.8 | 1.0 | 1.2 | 1.4 |
|--------------------------|-------|-------|-------|-------|-------|-------|-------|-------|
| OD value | 0.139 | 0.310 | 0.465 | 0.668 | 0.839 | 1.018 | 1.142 | 1.339 |
| | 0.140 | 0.292 | 0.489 | 0.669 | 0.821 | 1.040 | 1.176 | 1.317 |
| Average OD | 0.140 | 0.301 | 0.477 | 0.668 | 0.830 | 1.029 | 1.159 | 1.328 |
| Absoluted OD | 0.000 | 0.161 | 0.338 | 0.529 | 0.690 | 0.889 | 1.019 | 1.188 |



Appendix II Example Analysis


Example analysis:

Take 0.1 g of green pepper, treat the sample according to the sample preparation and carry the assay according to the operation steps. The results are as follows:

standard curve: $y = 0.8729x - 0.0112$, the average OD value of the sample is 0.368, the average OD value of the control is 0.247, and the calculation result is:

$$\begin{aligned}\alpha\text{-Amylase activity (U/g tissue)} &= (0.368 - 0.247 + 0.0112) \div 0.8729 \times 0.15 \div 5 \div 0.1 \times 10 \div 0.075 \\ &= 6.06 \text{ U/g tissue}\end{aligned}$$

Detect 1% green pepper tissue homogenate, 1% epipremnum aureum tissue homogenate, 1% corn grain tissue homogenate and 1% daucus carota tissue homogenate according to the protocol, the result is as follows:



Statement

1. This assay kit is for Research Use Only. We will not response for any arising problems or legal responsibilities causing by using the kit for clinical diagnosis or other purpose.
2. Please read the instructions carefully and adjust the instruments before the experiments. Please follow the instructions strictly during the experiments.
3. Protection methods must be taken by wearing lab coat and latex gloves.
4. If the concentration of substance is not within the detection range exactly, an extra dilution or concentration should be taken for the sample.
5. It is recommended to take a pre-test if your sample is not listed in the instruction book.
6. The experimental results are closely related to the situation of reagents, operations, environment and so on. Elabscience will guarantee the quality of the kits only, and NOT be responsible for the sample consumption caused by using the assay kits. It is better to calculate the possible usage of sample and reserve sufficient samples before use.

