
(FOR RESEARCH USE ONLY. DO NOT USE IT IN CLINICAL DIAGNOSIS !)

Human Hepatitis B Virus Surface (HBs) Antibodies ELISA Kit

Catalog No: GEH144

48T

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Tel: 400-967-3365

Email: techsupport@uni-science.com

Website: www.elabscience.cn

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Test principle

This ELISA kit uses Sandwich-ELISA as the method to detect the anti - HBs in human serum and plasma. The ELISA Microtiter plate provided in this kit has been pre-coated with purified Hepatitis B Virus Surface (HBs) antigen. Samples are added to the ELISA Microtiter plate wells and the HBs antibody in which will combine with the pre-coated antigen to form antigen-antibody compound. Free components are washed away. The HRP conjugated Mouse-anti-human antibody is added to each well and react with the compound to form “HBs antigen- HBs antibody-HRP antigen-” compound. The substrate reagent is added to initiate the color developing reaction. The presence of anti - HBs can be determined by measuring the absorbance value by using a microplate reader with 450 nm (630 nm) wavelength.

Kit components

Item	Specifications
ELISA Microtiter plate	48 wells
20×Concentrated Wash Buffer	15 mL
HRP Conjugate	3 mL
Substrate Reagent A	3 mL
Substrate Reagent B	3 mL
Stop Solution	3 mL
Positive Control	0.4 mL
Negative Control	0.4 mL
Plate Sealer	3 pieces
Scaled Bag	1 piece
Manual	1 copy

Note: All reagent bottle caps must be tightened to prevent evaporation and microbial pollution.

Other supplies required

Microplate reader with 450nm wavelength filter

High-precision transferpettor, EP tubes and disposable pipette tips

37°C Incubator

Deionized water

Absorbent paper

Notes

1. Please read the manual carefully before use, changes of operation may result in unreliable results.
2. Wear gloves and work clothes during experiment, and the disinfection and isolation system should be strictly executed. All the waste should be handled as contaminant.
3. The stop solution is corrosive, it should be avoided to contact with skin and clothing. Wash immediately with plenty of water if contact it carelessly.
4. The ELISA Microtiter plate obtained from cold storage conditions should be adjusted to room temperature before use. The unused plate should be kept in a sealed bag with desiccant.
5. Concentrated washing liquid at low temperature condition is easy to crystallization, it should be adjusted to room temperature in order to dissolve completely before use.
6. The results shall depend on the readings of the micro-plate Reader.
7. **Each reagent is optimized for use in the GEH144. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other GEH144 with different lot numbers.**
8. All the samples and waste material should be treated as infective material according to the relevant rules of biosafety.

Storage and expiry date

Store unopened at 2 to 8°C. Do not freeze.

Please store the opened plate at 2~8°C, the shelf life of the opened kit is up to 1 month.

Expiry date: expiration date is on the box.

Requirements of sample

1. **Serum/plasma:** Human serum and plasma can be used as detected sample. Fresh collected samples should be fully centrifuged, then take clear liquid for test. Suspended fibrous protein may cause a false positive.
2. Anticoagulant (EDTA, sodium citrate and heparin sodium) in samples do not affect the result of the experiment in general. Endogenous interference substances in serum such as blood lipids, cholestylin, hemoglobin, rheumatoid factors positive samples may not affect the results.
3. Samples can be stored at 2~8℃ for one week and stored at -20 ℃ for more than a week. Avoid freeze-thaw cycles. Freezing samples should be mixed fully before test.
4. **Wash Buffer:** The **20×Concentrated Wash Buffer** should be adjusted to room temperature to make the sediment dissolved fully before use, then dilute it with deionized water at 1:19.

Assay procedure

Restore all reagents and samples to room temperature (25℃) before use. All the reagents should be mixed thoroughly by gently swirling before pipetting. Avoid foaming. The unused ELISA Microtiter plate should be sealed as soon as possible and stored at 2-8℃.

1. **Number:** number the sample and standard in order (multiple well), and keep a record of standard wells and sample wells. Set 1 well for blank control, 2 wells for negative control and 1 well for positive control. **Samples need test in duplicate.**(Blank well is not necessary for dual-wavelength detection)
2. **Add sample:**
 - a) Add 50 μL of **negative control/positive control** to **negative control/positive control** wells, keep the blank control well empty.
 - b) Add 50 μL of serum sample to other sample wells.
 - c) Gently tap the plate to mix thoroughly.
3. **HRP conjugate:** add 50 μL of **HRP Conjugate** to each well except the blank control well.
4. **Incubate:** cover the ELISA Microtiter plate with sealer. Incubate for 60 minutes at 37℃ in shading light.
5. **Wash:** after incubation, remove the plate sealer and aspirate the liquid of each well. Repeat the washing procedure for 5 times with **Wash Buffer** and immerse for 30-60s each time. Invert the plate and pat it against thick clean absorbent paper (If bubbles exist in the wells, clean tips can be used to prick them).
6. **Add Substrate:** add 50 μL of **Substrate Reagent A** and 50 μL of **Substrate Reagent B** to each well. Gently tap the plate to mix thoroughly. Cover with a new plate sealer. Incubate for 15 minutes at 37℃ in shading light.
7. **Stop reaction:** add 50 μL of **Stop Solution** to each well, gently tap the plate to mix thoroughly.
8. **OD Measurement:** set the Microplate Reader wavelength at 450 nm (it is recommended to set the dual wavelength at 450 nm/630 nm) to detect A value of each well. Blank well is not essential when using dual wavelength 450 nm/630 nm for detection. **Note: Read the results within 30 min.**

Reference value

Normally, blank well (just substrate agent and stop solution) absorbance: $A_{450} \leq 0.08$;
positive control (PC): $A_{450} > 0.80$ and average A value of negative control (NC): $A_{450} < 0.08$.

Interpretation of test results

Use each test result independently. Determine the result according to the Cut Off value.

Cut Off(C.O) = $2.1 \times$ average A value of negative control(NC) (when average A_{450} of NC < 0.05 , calculate at 0.05; while average A_{450} of NC ≥ 0.05 , calculate at the actual value).

1. Positive result: average A value of sample \geq Cut Off.
2. Negative result: average A value of sample $<$ Cut Off.

Limitations

1. This kit is only for qualitative detection of Hepatitis B Virus Surface (HBs) Antibody in human.
2. The detection results of this kit are only for reference. For confirmation of the result, please combine the symptoms and other methods of detection, this detection cannot be used as the only criteria for result.