

## **DON (Deoxynivalenol) ELISA Kit**

Catalog No: E-TO-E023

96T/96T\*3

<b>Version Number:</b>	V1.1
<b>Replace version:</b>	V1.0
<b>Revision Date:</b>	2026.02.05

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Toll-free: 1-888-852-8623 Tel: 1-832-243-6086 Fax: 1-832-243-6017

Email: [techsupport@elabscience.com](mailto:techsupport@elabscience.com)

Website: [www.elabscience.com](http://www.elabscience.com)

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

## Test principle

This kit uses Competitive-ELISA as the method for the quantitative detection. It can detect Deoxynivalenol (DON) in samples, such as grain, feed, etc. This kit is composed of ELISA Microtiter plate, HRP conjugate, antibody working solution, standard and other supplementary reagents. The Microtiter plate provided in this kit has been pre-coated with coupled antigen. During the reaction, DON in the samples or standard competes with coupled antigen on the solid phase supporter for sites of anti-DON antibody. Then Horseradish Peroxidase (HRP) conjugate is added to each microtiter plate well, and substrate reagent is for color development. There is a negative correlation between the OD value of samples and the concentration of DON. The concentration of DON in the samples can be calculated by comparing the OD of the samples to the standard curve.

## Technical indicator

**Reaction mode** (Incubation time and temperature):  $25 \pm 2^{\circ}\text{C}$ ; 20min, 10-15min

**Detection limit:** Grain, Feed ---200 ppb

**Cross-reactivity:** Deoxynivalenol ---100%, 3-acetyldeoxynivalenol ---1000%

**Sample recovery rate:**  $100\% \pm 30\%$

## Kits components

Item	Specifications
ELISA Microtiter plate	96 wells
Standard Liquid	1.5 mL each (ppb=ng/mL=ng/g) (0 ppb, 5 ppb, 15 ppb, 45 ppb, 135 ppb, 405 ppb)
HRP Conjugate	7 mL
Antibody Working Solution	7 mL
Substrate Reagent A	7 mL
Substrate Reagent B	7 mL
Stop Solution	7 mL
20×Concentrated Wash Buffer	25 mL
Sample Diluent	100 mL
Plate Sealer	1 piece
Sealed Bag	1 piece
Manual	1 copy

Note: All reagent bottle caps must be tightened to prevent evaporation and microbial pollution.

## Other materials required but not supplied

**Instrument:** Microplate reader, Printer, Homogenizer, Vortex mixer, Centrifuge, Graduated pipette, Balance (sensitivity: 0.01g).

**High-precision transferpeltor:** Single channel (20-200  $\mu$ L, 100-1000  $\mu$ L), Multichannel (300  $\mu$ L).

**Reagents:** Methanol (Analytical reagent).

## Notes

1. The overall OD value will be lower when reagents have not been brought to room temperature before use or room temperature is below 25°C.
2. If the wells turn dry during the washing procedure, it will lead to bad linear standard curve and poor repeatability. Operate the next step immediately after wash.
3. Mix thoroughly and wash the plate completely. The consistency of wash procedure can strongly affect the reproducibility of this ELISA kit.
4. FOR RESEARCH USE ONLY. ELISA Microtiter plate should be covered by plate sealer. Avoid the kit to strong light.
5. **Each reagent is optimized for use in the E-TO-E023. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other E-TO-E023 with different lot numbers.**
6. Substrate Reagent should be abandoned if it turns blue color. When OD value of standard (concentration: 0) < 0.8 unit (A450nm < 0.8), it indicates the reagent be deteriorated.
7. Stop solution is caustic, avoid contact with skin and eyes.
8. As the OD values of the standard curve may vary according to the conditions of the actual assay performance (e.g. operator, pipetting technique, washing technique or temperature effects), the operator should establish a standard curve for each test.
9. Even the same operator might get different results in two separate experiments. In order to get reproducible results, the operation of every step in the assay should be controlled.
10. **If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.**
11. The kit is used for rapid screening of actual samples. If the test result is positive, the instrument method such as HPLC, LC/MS, etc. can be used for quantitative confirmation.

## Storage and expiry date

Store the kit at 2-8°C. Do not freeze any test kit components.

Return any unused microwells to their original foil bag and reseal them together with the desiccant provided and further store at 2-8 °C. After opening, the kit is stable for up to 1 month.

**Expiry date:** expiration date is on the packing box.

## Experimental preparation

Restore all reagents and samples to room temperature before use.

Open the microplate reader in advance, preheat the instrument, and set the testing parameters.

### 1. Sample pretreatment Notice:

Experimental apparatus should be clean and the pipette should be disposable to avoid cross-contamination during the experiment.

### 2. Solution preparation

*Please prepare solution according to the number of samples. Don't use up all components in the kit at once!*

Solution 1: **70% Methanol**.

**Methanol (V):** Deionized water (V) =7:3, mix fully.

Solution 2: **Wash Buffer**

Dilute the **20×Concentrated Wash Buffer** with deionized water. (20×Concentrated Wash Buffer (V): Deionized water (V) =1:19)

### 3. Sample pretreatment procedure.

#### 3.1 Pretreatment of grain (beans, rice, corn, oats, wheat), feed sample:

- (1) Homogenize the representative dry sample with a homogenizer and mix fully
- (2) Weigh  $5\pm 0.05$  g of homogenate sample in to 50 mL centrifuge tube, add 25mL of **70% Methanol** (Solution 1), vortex for 5 min, centrifuge at 4000 rpm for 5 min at room temperature;
- (3) Take 100 $\mu$ L of supernatant to another centrifuge tube, add 700  $\mu$ L of **Sample Diluent**, mix fully;
- (4) Take 50  $\mu$ L for analysis.

**Note: Sample dilution factor: 40, detection limit: 200 ppb**

## Assay procedure

Restore all reagents and samples to room temperature before use. All the reagents should be mixed thoroughly by gently swirling before pipetting. Avoid foaming. The unused ELISA Microtiter plate should be sealed as soon as possible and stored at 2-8°C.

1. **Number:** number the sample and standard in order (multiple well), and keep a record of standard wells and sample wells. **Standard and Samples need test in duplicate.**
2. **Add sample:** add 50 µL of **Standard or Sample** per well, then add 50 µL of **HRP Conjugate** to each well, then add 50 µL of **Antibody Working Solution**, cover the plate sealer, oscillate for 10 s gently to mix thoroughly, incubate at  $25 \pm 2^\circ\text{C}$  for 20 min in shading light.
3. **Wash:** uncover the sealer carefully, remove the liquid of each well. Immediately add 260 µL of **Wash Buffer** (Solution 2) to each well and wash. Repeat wash procedure for 4 times, 15-30 s intervals/time. Invert the plate and pat it against thick clean with absorbent paper (If bubbles exist in the wells, clean tips can be used to prick them).
4. **Color Development:** add 100 µL of Substrate mixed solution to each well (**Substrate Reagent A** and **Substrate Reagent B** are fully mixed at ratio 1:1 by volume, the mixture should be used within 5 min, avoid using metal containers or stirring the reagents). Gently oscillate for 15 s to mix thoroughly. Incubate at  $25 \pm 2^\circ\text{C}$  for 10-15 min in shading light. (The reaction time can be extended according to the actual color change).
5. **Stop reaction:** add 50 µL of **Stop Solution** to each well. Gently oscillate to mix thoroughly.
6. **OD Measurement:** determine the optical density (OD value) of each well at 450 nm (reference wavelength 630 nm) with a microplate reader. This step should be finished in 5 min after stop reaction.

## Result analysis

1. **Absorbance (%)**= $A/A_0 \times 100\%$

A: Average absorbance of standard or sample

$A_0$ : Average absorbance of 0 ppb Standard

2. **Drawing and calculation of standard curve:**

Create a standard curve by plotting the absorbance percentage of each standard on the y-axis against the log concentration on the x-axis to draw a semi-logarithmic plot. Add average absorbance value of sample to standard curve to get corresponding concentration. **If samples have been diluted, the concentration calculated from the standard curve must be multiplied by the dilution factor.**

For this kit, it is more convenient to use professional analysis form for accurate and fast analysis on a large number of samples.

### Deoxynivalenol (E-TO-E023) Standard Curve

