

EasySort™ Human CD3⁺T Cell Isolation Kit

Cat. No: MIH001N

Size: 10 Assays/100 Assays/200 Assays

| Component | Component Name | 10 Assays | 100 Assays | 200 Assays | Storage |
|-----------|--|-----------|------------|------------|---------|
| MIH001NA | EasySort™ Human CD3 ⁺ T Beads Streptavidin 1.0-N | 160 µL | 800 µL×2 | 800 µL×4 | 2-8°C |
| MIH001NB | EasySort™ Human CD3 ⁺ T Cell Isolation Cocktail | 120 µL | 1.2 mL | 1.2 mL×2 | 2-8°C |
| | Manual | | | 1 copy | |

Storage

Store at 2-8°C with shading light for 1 year. Avoid freezing and thawing.

Description

The EasySort™ Human CD3⁺T cell isolation kit is a product that enables rapid and simple isolation of high-purity human CD3⁺T cells. This kit uses a negative selection method and is suitable for isolating CD3⁺T from human PBMC sample. Different biotinylated monoclonal antibodies are used to label non-target cells (non-human CD3⁺T cells). Subsequently, streptavidin-conjugated magnetic beads are employed to deplete these non-target cells, thereby obtaining highly purified human CD3⁺T cells. The isolated human CD3⁺T cells are free of any antibodies and magnetic bead labels, remain in an unstimulated, naïve state, and are ready for direct use in downstream applications.

The EasySort™ Human CD3⁺T Cell Isolation Kit has been tested by magnetic cell separation followed by flow cytometric analysis of cells from fresh human PBMC sample. An assay is defined as 12 µL antibody and 16 µL beads to be used to isolate 1×10⁷ cells.

Reagents and Materials Not Supplied

1. Reagents:

Phosphate buffered saline (PBS), fetal bovine serum (FBS), EDTA, Human peripheral blood mononuclear cells separation solution, DNase I

2. Materials:

70 µm mesh nylon strainer, 1.5 mL/2 mL EP tube, 15 mL/50 mL centrifuge tube, flow tube

3. Instrument:

Optical microscope, horizontal centrifuge, magnetic rack

For Research Use Only

Experimental Operation

NOTE: The following operations must be performed under sterile conditions

➤ Isolation buffer preparation

Add fetal bovine serum (final concentration of 2%) and EDTA (final concentration of 2 mM) to PBS buffer and filter the prepared buffer with 0.22 µm filter.

NOTE: Sealed store the prepared buffer at 4°C and use within 1 week. In addition, 2% fetal bovine serum can be replaced by 0.5% BSA.

➤ Sample Preparation and Processing

1. Fresh human PBMC: PBMC sample is obtained from fresh human whole blood by density gradient centrifugation. Wash PBMC twice with isolation buffer, centrifuge at 300 g for 5 min, filter the PBMC through a 70 µm mesh nylon strainer and adjust the cell density to 1×10^8 cells/mL for cell isolation.
2. Frozen PBMC: The frozen PBMC should be incubated with DNase I solution (PBS) at a concentration of 100 µg/mL for 15 min at room temperature before cell isolation. Wash sample twice with isolation buffer, centrifuged at 300 g for 5 min. Filter aggregated suspensions through a 70 µm mesh nylon strainer and adjust cell density at 1×10^8 cells/mL.

Note: Generally, approximately 1×10^7 PBMC cells can be obtained from 10 mL of human blood. After preparing a single-cell suspension from fresh human blood, perform the cell isolation experiment within 1-2 hours, as a longer interval will affect the final isolated cell purity and cell viability.

➤ Cell Isolation

- a) Prepare 100 µL of cell suspension (about 1×10^7 cells), add 12 µL Human CD3⁺T Cell Isolation Cocktail, gently pipette up and down 6-8 times with a pipette to mix, then incubate for 5 min at room temperature.

Note: Please ensure that the cells are in a single-cell suspension. Before sample dilution, filter the samples through a 70 µm cell sieve. The frozen PBMC sample need to be treated with DNase I and then filtered through a 70 µm cell sieve again before isolating.

- b) Add isolation buffer to a final volume of 2 mL, centrifuge at 300 g for 5 min. Discard the supernatant, and then resuspend the cells with 100 µL isolation buffer.

Note:

- If the total volume of the cell suspension exceeds 1 mL, the volume of the added isolation buffer shall be no less than the total volume of the cell suspension. For example, if the total volume of the cell suspension is 1.5 mL, the volume of the isolation buffer added shall be ≥ 1.5 mL.
- To maintain consistent cell density, the volume of cell isolation buffer for cell resuspension shall be identical to that of the input cell suspension. In the protocol example, if 100 µL of cell suspension is used as the starting input, cells should be

For Research Use Only

resuspended with an equal volume of 100 μ L cell isolation buffer.

- c) Wash Human CD3⁺T Beads Streptavidin 1.0-N: Place a clean flow cytometry tube or a centrifuge tube compatible with the magnetic rack into a tube rack. Pipette 1 mL of isolation buffer into the tube, then add 16 μ L of magnetic beads directly into the aforementioned 1 mL of isolation buffer. Mix by pipetting up and down 6-8 times. Place the flow cytometry tube or centrifuge tube on a magnetic rack (provided by the user) and magnetically separate at room temperature for 5 min. At this point, the magnetic beads are attracted to the tube wall. Keep the tube on the magnetic rack, discard the supernatant, and then remove the tube from the magnetic rack.

Note: If the total volume of magnetic beads to be washed is greater than 1 mL, use a 1:1 volume ratio of isolation buffer to beads during the washing step.

- d) Resuspend the magnetic beads using the cell suspension from step b): Aspirate the cell suspension and pipette the beads off the tube wall to the bottom of the tube (Note: avoid generating bubbles). Mix by pipetting up and down 6-8 times, then incubate at room temperature for 5 min.

Note:


- If more than 1×10^7 cells are to be isolated, increase the amount of Human CD3⁺T Cell Isolation Cocktail and Human CD3⁺T Beads Streptavidin 1.0-N proportionally while ensuring the cell density remains 1×10^8 cells/mL. If fewer than 1×10^7 cells are to be isolated, resuspend the cells with 100 μ L isolation buffer, add 12 μ L Human CD3⁺T Cell Isolation Cocktail and 16 μ L washed Human CD3⁺T Beads Streptavidin 1.0-N.
- The 5 mL flow tube is suitable for isolation of cell suspension ≤ 1 mL (2×10^8 cells). 10 mL or 15 mL centrifuge tube is suitable for isolation of cell suspension ≤ 4 mL (8×10^8 cells).

- e) Add isolation buffer to a final volume of 2.5 mL, (If the volume of the cell suspension for isolation is >1 mL, resuspend in an equal volume of isolation buffer), mix gently with a pipette by blowing up and down for 6-8 times until no particles of magnetic beads are visible. Put the tube on a 5 mL magnetic rack (self-provided) and stand for 5 min.

Note: Please mix the liquid thoroughly to avoid the magnetic beads clumping and affecting the isolation efficiency.

- f) Transfer the cell suspension to a clean centrifuge tube, centrifuge at 300 g for 5 min. Discard the supernatant, resuspend the cells with buffer required for the subsequent experiments.

Typical data



The CD3⁺T cells isolated from normal PBMC were stained with PE Anti-Human CD3 Antibody [OKT3] (E-AB-F1001D). The purities of the start and final isolated fractions were 60.92% and 96.73%, respectively.

Cautions

1. This kit is for research use only.
2. Please take safety precautions and follow the procedures of laboratory reagent operation.
3. All components of the kit should be stored at 2-8°C and protected from freezing and thawing.
4. Sample type, sample preparation and experimental operation have an important impact on the final isolated cell purity.
5. The quality of pre-isolated PBMC sample is critically impacts the separation efficiency of this product. It is recommended to test whether the percentage of CD3⁺T cells is in the normal physiological range (45%-70%) after the preparation of PBMC sample. It is recommended to re-prepare the PBMC sample when percentage of target cell population is lower than its normal distribution.
6. Single-cell suspension for cell isolation shall be filtered through a cell strainer to remove cell clumps and tissue debris, preventing cell aggregation from compromising isolation purity.
7. Perform isolation immediately after preparing the cell suspension, as cell viability will decrease with longer storage time.
8. When adding the antibody cocktail and aspirating the magnetic beads for washing, pipette them directly to the bottom of the tube to avoid adhesion to the wall, which would result in loss of components.
9. In order to ensure the activity of the cells, the whole process of the experiment should be completed on ice as much as possible, except for the incubation at room temperature.
10. It is recommended to use low adsorption pipette tips and centrifuge tubes to avoid the loss of magnetic beads and antibodies due to adsorption.
11. The kit should be used in combination with a magnetic rack.

For Research Use Only