

FQNs (Fluoroquinolones) ELISA Kit

Catalog No: E-FS-E054

96T/96T*3

Version Number:	V1.2
Replace version:	V1.1
Revision Date:	2024.03.14

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Toll-free: 1-888-852-8623 Tel: 1-832-243-6086 Fax: 1-832-243-6017

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Test principle

This kit uses Competitive-ELISA as the method for the quantitative detection. It can detect Fluoroquinolones (FQN_s) in samples, such as milk, muscle, honey, etc. This kit is composed of ELISA Microtiter plate, HRP conjugate, antibody working solution, standard and other supplementary reagents. The microtiter plate in this kit has been pre-coated with coupled antigen. During the reaction, FQN_s in the samples or standard competes with coupled antigen on the solid phase supporter for sites of anti-FQN_s antibody. Then Horseradish Peroxidase (HRP) conjugate is added to each microtiter plate well, and substrate reagent is added for color development. There is a negative correlation between the OD value of samples and the concentration of FQN_s. The concentration of FQN_s in the samples can be calculated by comparing the OD of the samples to the standard curve.

Technical indicator

Reaction mode(Incubation time and temperature): 25°C;45 min, 15min.

Detection limit: Muscle ---0.6 ppb; Honey ---0.8 ppb; Milk, Eggs ---6 ppb; Milk powder---12 ppb; Urine---1 ppb.

Cross-reactivity:

Enrofloxacin	100%
Norfloxacin	>100%
Ciprofloxacin	>100%
Lomefloxacin	>100%
Flumequine	>100%
Peflacin	>100%
Danofloxacin	>100%
Sarafloxacin	>100%
Difloxacin	84%
Enoxacin	66%
Ofloxacin (racemate)	58%
Oxolinic acid	28%
Levofloxacin	10%
Marbofloxacin	4%

Sample recovery rate: Muscle, Honey, Milk, Milk powder, Eggs, Urine ---85% ± 15%.

Kits components

Item	Specifications
ELISA Microtiter plate	96 wells
Standard Liquid	1 mL each (ppb=ng/mL=ng/g) (0 ppb, 0.2 ppb, 0.6 ppb, 1.8 ppb, 5.4 ppb, 16.2 ppb)
HRP Conjugate	5.5 mL
Antibody Working Solution	5.5 mL
Substrate Reagent A	6 mL
Substrate Reagent B	6 mL
Stop Solution	6 mL
20×Concentrated Wash Buffer	40 mL
5×Reconstitution Buffer	50 mL
Plate Sealer	3 pieces
Sealed Bag	1 piece
Manual	1 copy

Note: All reagent bottle caps must be tightened to prevent evaporation and microbial pollution.

Other materials required but not supplied

Instrument: Microplate reader, Printer, Homogenizer, Vortex mixer, Nitrogen evaporators, Water bath, Centrifuge, Graduated pipette, Balance (sensitivity 0.01g).

Micropipette: Single channel (20-200 μ L, 100-1000 μ L), Multichannel (30-300 μ L).

Reagents: Acetonitrile, N-hexane, HCl, Dichloromethane.

Notes

1. The overall OD value will be lower when reagents have not been brought to room temperature before use or room temperature is below 25°C.
2. If the wells turn dry during the washing procedure, it will lead to bad linear standard curve and poor repeatability. Operate the next step immediately after wash.
3. Mix thoroughly and wash the plate completely. The consistency of wash procedure can strongly affect the reproducibility of this ELISA kit.
4. FOR RESEARCH USE ONLY. ELISA Microtiter plate should be covered by plate sealer. Avoid the kit to strong light.
5. **Each reagent is optimized for use in the E-FS-E054. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other E-FS-E054 with different lot numbers.**
6. Substrate Reagent should be abandoned if it turns blue color. When OD value of standard (concentration: 0) < 0.5 unit (A450nm < 0.5), it indicates the reagent may be deteriorated.
7. Stop solution is caustic, avoid contact with skin and eyes.
8. As the OD values of the standard curve may vary according to the conditions of the actual assay performance (e.g. operator, pipetting technique, washing technique or temperature effects), the operator should establish a standard curve for each test.
9. Even the same operator might get different results in two separate experiments. In order to get reproducible results, the operation of every step in the assay should be controlled.
10. **For mentioned sample fast and efficient extraction methods are included in the kit description. Please consult technical support for the applicability if other sample need to be tested.**
11. The kit is used for rapid screening of actual samples. If the test result is positive, the instrument method such as HPLC, LC/MS, etc. can be used for quantitative confirmation.

Storage and expiry date

Store the kit at 2-8°C. Do not freeze any test kit components.

Return any unused microwells to their original foil bag and reseal them together with the desiccant provided and further store at 2-8°C.

Expiry date: expiration date is on the packing box.

Experimental preparation

Restore all reagents and samples to room temperature before use.

Open the microplate reader in advance, preheat the instrument, and set the testing parameters.

1. Sample pretreatment Notice:

Experimental apparatus should be clean, and the pipette should be disposable to avoid cross-contamination during the experiment.

2. Solution preparation

Please prepare solution according to the number of samples. Don't use up all components in the kit at once!

Solution 1: 0.15 M HCl Solution

Dilute 5 mL of **HCl** to 400 mL with deionized water, mix fully.

Solution 2: Sample Extraction Solution (for livestock, fish, shrimp, honey sample)

Add 10 mL of **0.15 M HCl Solution** (Solution 1) to 90 mL of **Acetonitrile**, mix fully

Solution 3: Reconstitution Buffer

Dilute the **5×Reconstitution Buffer** with deionized water. (5×Reconstitution Buffer(V) :Deionized water(V)=1:4). The Reconstitution buffer can be store at 4°C for a month.

Solution 4: Wash Buffer

Dilute **20×Concentrated Wash Buffer** with deionized water. (20×Concentrated Wash Buffer (V): Deionized water (V) = 1:19).

3. Sample pretreatment procedure

3.1 Pretreatment of muscle (livestock, fish, shrimp) sample:

- (1) Remove fat from sample. Homogenize the representative sample with a homogenizer and mix fully.
- (2) Weigh 2 ± 0.05 g of homogenate sample into a 50 mL centrifuge tube.
- (3) Add 8 mL of **Sample Extraction Solution** (Solution 2) and vortex for 5min. Centrifuge at 4000 rpm for 10 min at room temperature.
- (4) Take 2 mL of the clear upper organic layer solution to a 10 mL of clean and dry glass tube, dry at 50-60°C with nitrogen evaporators or water bath. (Please do it in a ventilated environment.)
- (5) Add 1 mL of **N-hexane** and vortex for 2 min. Then add 1 mL of **Reconstitution Buffer** (Solution 3) and vortex for 30 s to mix fully. Centrifuge for 5 min at 4000 rpm at room temperature.
- (6) Discard the N-hexane upper layer, take 50 μ L of the lower layer solution for analysis.

Note: Sample dilution factor: 2, detection limit: 0.6 ppb

3.2 Pretreatment of honey sample:

- (1) Weigh 1 ± 0.05 g of honey into a 50 mL centrifuge tube, add 6 mL of **Sample Extraction Solution** (Solution 2) and vortex for 5 min to ensure thoroughly dissolved.
- (2) Add 3 mL of **Reconstitution Buffer** (Solution 3) and 11 mL of **Dichloromethane**, vortex for 5 min. Then centrifuge at 4000 rpm for 5 min at room temperature.
- (3) Remove the supernatant and transfer 8 mL of the upper layer organic solution to another tube. Dry at 50-60°C with nitrogen evaporators or water bath.
- (4) Dissolve the residue with 1 mL of **Reconstitution Buffer** (Solution 3). Add 1 mL of **N-hexane** and vortex for 30 s. Centrifuge for 5 min at 4000 rpm at room temperature.
- (5) Discard the N-hexane upper layer, take 50 μ L of the lower layer solution for analysis.

Note: Sample dilution factor: 2, detection limit: 0.8 ppb

3.3 Pretreatment of milk sample:

- (1) Dilute 25 μL of milk with 475 μL of **Reconstitution Buffer** (Solution 3), vortex for 1 min to mix fully.
- (2) Take 50 μL for analysis.

Note: Sample dilution factor: 20, detection limit: 6 ppb

3.4 Pretreatment of milk powder sample:

- (1) Weigh 0.5 ± 0.05 g of homogenate sample into 10 mL centrifuge tube, add 5 mL of deionized water and vortex to dissolve fully.
- (2) Mix 100 μL of sample solution with 400 μL of **Reconstitution Buffer** (Solution 3) and vortex for 1 min.
- (3) Take 50 μL for analysis.

Note: Sample dilution factor: 50, detection limit: 12 ppb

3.5 Pretreatment of eggs sample:

- (1) Weigh 1 ± 0.05 of homogenate egg into 10 mL centrifuge tube, add 5 mL of deionized water and vortex to dissolve fully.
- (2) Mix 100 μL of sample solution with 400 μL of **Reconstitution Buffer** (Solution 3) and vortex for 1 min.
- (3) Take 50 μL for analysis.

Note: Sample dilution factor: 30, detection limit: 6 ppb

3.6 Pretreatment of urine (swine) sample:

- (1) Take 1 mL of clear urine sample (If the urine sample is turbid, it should be filtered or 4000 rpm for 5 minutes until the urine sample is clear, take the supernatant for testing) and add 4 mL of **Reconstitution Buffer** (Solution 3). Vortex for 30 s.
- (2) Take 50 μL for analysis.

Note: Sample dilution factor: 5, detection limit: 1 ppb

Assay procedure

Restore all reagents and samples to room temperature (25°C) before use. All the reagents should be mixed thoroughly by gently swirling before pipetting. Avoid foaming. The unused ELISA Microtiter plate should be sealed as soon as possible and stored at 2-8°C.

1. **Number:** Number the sample and standard in order (multiple well), and keep a record of standard wells and sample wells. **Standard and Samples need test in duplicate.**
2. **Add Sample:** Add 50 µL of **Standard** or **Sample** per well, then add 50 µL of **HRP Conjugate** to each well. Add 50 µL of **Antibody Working Solution**. Gently oscillate for 5 s to mix thoroughly and cover the plate with sealer. Incubate at 25°C for 45 min in shading light.
3. **Wash:** Uncover the sealer carefully, remove the liquid in each well. Immediately add 300 µL of **Wash Buffer** (Solution 4) to each well and wash. Repeat the wash procedure for 5 times, 30 s intervals/time. Invert the plate and pat it against thick clean absorbent paper (If bubbles exist in the wells, clean tips can be used to prick them).
4. **Color Development:** Add 50 µL of **Substrate Reagent A** to each well, and then add 50 µL of **Substrate Reagent B**. Gently oscillate for 5 s to mix thoroughly. Incubate at 25°C for 15 min in shading light (The reaction time can be extended according to the actual color change).
5. **Stop Reaction:** Add 50 µL of **Stop Solution** to each well, oscillate gently to mix thoroughly.
6. **OD Measurement:** determine the optical density (OD value) of each well at 450 nm (reference wavelength 630 nm) with a microplate reader. This step should be finished in 10 min after stop reaction.

Result analysis

1. Absorbance(%)= $A/A_0 \times 100\%$

A: Average absorbance of standard or sample

A_0 : Average absorbance of 0 ppb Standard

2. Drawing and calculation of standard curve

Create a standard curve by plotting the absorbance percentage of each standard on the y-axis against the log concentration on the x-axis to draw a semi-logarithmic plot. Add average absorbance value of sample to standard curve to get corresponding concentration. **If samples have been diluted, the concentration calculated from the standard curve must be multiplied by the dilution factor.**

For this kit, it is more convenient to use professional analysis form for accurate and fast analysis of batch samples.

