(FOR RESEARCH USE ONLY. DO NOT USE IT IN CLINICAL DIAGNOSIS !)

Catalog No: E-BC-K781-M Specification: 48T(32 samples)/96T(80 samples) Measuring instrument: Microplate reader (405-415 nm) Detection range: 0.383-16.588 U/L

Elabscience[®] Phospholipase C (PLC) Activity Colorimetric Assay Kit

This manual must be read attentively and completely before using this product. If you have any problem, please contact our Technical Service Center for help:

Toll-free: 1-888-852-8623 Tell: 1-832-243-6086 Fax: 1-832-243-6017 Email: techsupport@elabscience.com Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Table of contents

Assay summary	3
Intended use	4
Detection principle	4
Kit components & storage	4
Materials prepared by users	5
Reagent preparation	5
Operating steps	8
Calculation	9
Appendix I Performance Characteristics	
Appendix П Example Analysis	
Statement	



Intended use

This kit can be used to detect the phospholipase C (PLC) activity in serum, plasma, cell, animal and plant tissue samples.

Detection principle

Phospholipase C (PLC) is a hydrolase that breaks down phosphatidylinositol diphosphate (PIP2) to form diacylglycerol (DAG) and inositol triphosphate (IP3). PLCS are widely found in bacteria, plants and animals. Its abnormal expression is associated with a variety of cancers, and studying the PLC mechanism can help develop new anticancer therapies. PLC also play a key role in nerve signaling and protection, helping to understand and treat neurodegenerative diseases. In addition, PLC plays an important role in insulin signaling and lipid metabolism, making it a hot topic in diabetes and obesity research.

PLC catalyzes the substrate to produce chromogenic substance has maximum absorption at 410 nm, and the PLC enzyme activity is calculated by measuring OD value at 410 nm.

Item	Component	Size 1(48 T)	Size 2(96 T)	Storage		
Reagent 1	Buffer Solution	10 mL ×1 vial	$20 \text{ mL} \times 1 \text{ vial}$	-20 °C, 12 months		
Reagent 2	Substrate	$0.75 \text{ mL} \times 1 \text{ vial}$	$1.5 \text{ mL} \times 1 \text{ vial}$	-20 °C, 12 months, shading light		
Reagent 3	5 mmol/L Standard Solution	$1 \text{ mL} \times 1 \text{ vial}$	$1 \text{ mL} \times 1 \text{ vial}$	-20 ℃, 12 months, shading light		
	Microplate	96 w	ells	No requirement		
	Plate Sealer	ces				

Kit components & storage

Note: The reagents must be stored strictly according to the preservation conditions in the above table. The reagents in different kits cannot be mixed with each other. For a small volume of reagents, please centrifuge before use, so as not to obtain sufficient amount of reagents.

Materials prepared by users

Instruments:

Microplate reader (405-415 nm, optimum wavelength: 410 nm), Incubator (37°C)

Reagents:

Normal saline (0.9% NaCl)

Reagent preparation

- ① Equilibrate all the reagents to 25°C before use.
- ② The preparation of 0.5 mmol/L standard solution:

Before testing, please prepare sufficient 0.5 mmol/L standard solution. For example, prepare 1000 μ L of 0.5 mmol/L standard solution (mix well 900 μ L of double distilled water and 100 μ L of 5 mmol/L standard solution). Keep it on ice during use protected from light and used up within same day.

 \bigcirc The preparation of working solution:

For each well, prepare 120 μ L of working solution (mix well 108 μ L of buffer solution and 12 μ L of substrate). Keep it on ice during use protected from light and used up within same day.

④ The preparation of standard curve:

Always prepare a fresh set of standards. Discard working standard dilutions after use.

Dilute 0.5 mmol/L standard solution with double distilled water diluent to a serial concentration. The recommended dilution gradient is as follows: 0, 0.10, 0.15, 0.20, 0.30, 0.35, 0.40, 0.50 mmol/L. Reference is as follows:

Item	1	2	3	4	5	6	\bigcirc	8
Concentration (mmol/L)	0	0.10	0.15	0.20	0.30	0.35	0.40	0.50
0.5 mmol/L standard (µL)	0	40	60	80	120	140	160	200
Double distilled water (µL)	200	160	140	120	80	60	40	0

Sample preparation

① Sample preparation:

Serum (plasma) samples: detect directly.

Tissue samples:

- Harvest the amount of tissue needed for each assay (initial recommendation 20 mg).
- ② Wash tissue in normal saline (0.9% NaCl).
- ③ Homogenize 20 mg tissue in 180 µL normal saline (0.9% NaCl) with a dounce homogenizer at 4°C.
- ④ Centrifuge at 10000×g for 10 min at 4°C to remove insoluble material. Collect supernatant and keep it on ice for detection.
- (5) Meanwhile, determine the protein concentration of supernatant (Animal tissue: E-BC-K318-M. Plant tissue: E-BC-K168-M).

Cell (adherent or suspension) samples:

- (1) Harvest the number of cells needed for each assay (initial recommendation 1×10^{6} cells).
- ② Wash cells with normal saline (0.9% NaCl).
- (3) Homogenize 1×10^{6} cells in 200 µL normal saline (0.9% NaCl) with a ultrasonic cell disruptor at 4°C.
- ④ Centrifuge at 10000 × g for 10 min at 4°C to remove insoluble material. Collect supernatant and keep it on ice for detection.
- (5) Meanwhile, determine the protein concentration of supernatant (E-BC-K318-M).

② Dilution of sample

The recommended dilution factor for different samples is as follows (for reference only):

Sample type	Dilution factor
Mouse serum	2-5
Rat serum	2-5
Human serum	2-5
10% Mouse liver tissue homogenate	1
10% Mouse kidney tissue homogenate	2-5
10% Mouse heart tissue homogenate	1
10% Mouse lung tissue homogenate	1
10% Mung bean sprout tissue homogenate	2-5
10% Edamame tissue homogenate	2-5
10% Peanut tissue homogenate	2-5
10% Broccoli tissue homogenate	2-5
2×10^6 HL-60 cells	1
1.8×10^6 Hela cells	1
2×10^6 CHO cells	1

Note: The diluent is normal saline (0.9% NaCl). For the dilution of other sample types, please do pretest to confirm the dilution factor.

Operating steps

(1) Standard wells: Add 20 μ L of standard solution with different concentrations to the corresponding wells.

Sample wells: Add 20 µL of sample to the corresponding wells.

- 2 Add 120 μL of working solution to each well.
- ③ Mix fully with microplate reader for 5 s and measure the OD value of sample well at 410 nm, as A₁. Incubate at 37 °C for 30 min and measure the OD value of each well at 410 nm, as A₂. (The standard curve is fitted to the standard well in A₂ value.)

Calculation

The standard curve:

1. Average the duplicate reading for each standard.

2. Subtract the mean A_2 value of the blank (Standard #1) from all standard readings. This is the absoluted A_2 value.

3. Plot the standard curve by using A₂ value of standard and correspondent concentration as y-axis and x-axis respectively. Create the standard curve (y = ax + b) with graph software (or EXCEL).

The sample:

1. Serum (plasma) samples:

Definition: The amount of enzyme in 1 L serum or plasma per 1 min that produce 1 μ mol product at 37 °C is defined as 1 unit.

$$\frac{\text{PLC activity}}{(\text{U/L})} = (\Delta A_{410} - b) \div a \div T \times f \times 1000$$

2. Tissue and cell samples:

Definition: The amount of enzyme in 1 g tissue or cell protein per 1 min that produce 1 μ mol product at 37 °C is defined as 1 unit.

$$\frac{\text{PLC activity}}{(\text{U/gprot})} = (\Delta A_{410} - b) \div a \div T \times f \div C_{\text{pr}} \times 1000$$

[Note]

 ΔA_{410} : $\Delta A_{410} = A_2 - A_1$.

T: Reaction time, 30 min.

f: Dilution factor of sample before test.

C_{pr}: Concentration of protein in sample, gprot/L.

1000: Unit conversion, 1 mmol/L= 1000 µmol/L.

Appendix I Performance Characteristics

1. Parameter:

Intra-assay Precision

Three human serum samples were assayed in replicates of 20 to determine precision within an assay. (CV = Coefficient of Variation)

Parameters	Sample 1	Sample 2	Sample 3		
Mean (U/L)	5.00	10.00	15.00		
%CV	1.8	2.0	3.0		

Inter-assay Precision

Three human serum samples were assayed 20 times in duplicate by three operators

Parameters	Sample 1	Sample 2	Sample 3		
Mean (U/L)	5.00	10.00	15.00		
%CV	6.7	7.8	9.9		

to determine precision between assays.

Recovery

Take three samples of high concentration, middle concentration and low concentration to test the samples of each concentration for 6 times parallelly to get the average recovery rate of 100.3%.

	Standard 1	Standard 2	Standard 3
Expected Conc. (U/L)	5	10	15
Observed Conc. (U/L)	5.0	10.0	15.2
Recovery rate (%)	100	100	101

Sensitivity

The analytical sensitivity of the assay is 0.383 U/L. This was determined by adding two standard deviations to the mean O.D. obtained when the zero standard was assayed 20 times, and calculating the corresponding concentration.

2. Standard curve

As the OD value of the standard curve may vary according to the conditions of the actual assay performance (e.g. operator, pipetting technique or temperature effects),

Concentration (mmol/L)	0	0.10	0.15	0.20	0.30	0.35	0.40	0.50
OD reduce	0.101	0.175	0.216	0.255	0.330	0.372	0.406	0.479
OD value	0.109	0.178	0.213	0.252	0.329	0.364	0.406	0.474
Average OD value	0.105	0.177	0.215	0.254	0.330	0.368	0.406	0.477
Absoluted OD value	0	0.072	0.110	0.149	0.225	0.263	0.301	0.372

so 1	the	stand	lard	curve	and	data	are	provi	ded	l as	bel	low	for 1	refe	erence	onl	y.
------	-----	-------	------	-------	-----	------	-----	-------	-----	------	-----	-----	-------	------	--------	-----	----



Appendix II Example Analysis

Example analysis:

Take 20 μ L of rat serum which dilute for 2 times and carry the assay according to the operation steps. The results are as follows:

Standard curve: y = 0.751 x - 0.0016, the A₁ of the sample well is 0.417, the A₂ of the sample well is 0.585, $\Delta A_{410} = A_2 - A_1 = 0.585 - 0.471 = 0.114$, and the calculation result is:

PLC activity (U/L) = (0.114 + 0.0016) $\div 0.751 \div 30 \times 2 \times 1000 = 10.26$ U/L

Detect rat serum (dilute for 2 times), 10% mouse kidney tissue homogenate (the concentration of protein is 9.635 gprot/L, dilute for 2 times), 10% edamame tissue homogenate (the concentration of protein is 3.201 gprot/L, dilute for 2 times) and 1.8×10^{6} HL-60 cells (the concentration of protein is 1.444 gprot/L) according to the protocol, the result is as follows:



Statement

- 1. This assay kit is for Research Use Only. We will not response for any arising problems or legal responsibilities causing by using the kit for clinical diagnosis or other purpose.
- 2. Please read the instructions carefully and adjust the instruments before the experiments. Please follow the instructions strictly during the experiments.
- 3. Protection methods must be taken by wearing lab coat and latex gloves.
- 4. If the concentration of substance is not within the detection range exactly, an extra dilution or concentration should be taken for the sample.
- 5. It is recommended to take a pre-test if your sample is not listed in the instruction book.
- 6. The experimental results are closely related to the situation of reagents, operations, environment and so on. Elabscience will guarantee the quality of the kits only, and NOT be responsible for the sample consumption caused by using the assay kits. It is better to calculate the possible usage of sample and reserve sufficient samples before use.